


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052919\
 Data File : BE099771.D
 Acq On : 29 May 2019 11:56
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE052019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 20 19:29:40 2019



AutoFind: Scans 2565, 2566, 2567; Background Corrected with Scan 2558

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.4	4421	PASS
68	69	0.00	2	1.5	94	PASS
69	198	0.00	100	29.6	6418	PASS
70	69	0.00	2	0.2	15	PASS
127	198	10	80	33.0	7152	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	21664	PASS
199	198	5	9	6.8	1478	PASS
275	198	10	60	30.3	6573	PASS
365	198	1	100	3.6	789	PASS
441	198	0.01	100	13.9	3008	PASS
442	198	50	100	79.9	17318	PASS
443	442	15	24	19.4	3367	PASS