

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060319\
 Data File : BE099790.D
 Acq On : 3 Jun 2019 19:17
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 04 03:55:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.592	0.573	3.2	101	-0.01
3	n-Nitrosodimethylamine	0.327	0.293	10.4	101	-0.01
4 S	2-Fluorophenol	1.011	1.098	-8.6	124	0.01
5 S	Phenol-d6	1.338	1.443	-7.8	129	0.01
6	bis(2-Chloroethyl)ether	1.190	1.218	-2.4	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.01
8 S	Nitrobenzene-d5	0.346	0.350	-1.2	127	0.01
9	Naphthalene	4.278	4.247	0.7	120	0.01
10	Hexachlorobutadiene	0.315	0.313	0.6	121	0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.740	-7.2	130	0.00
12	2-Methylnaphthalene	0.697	0.749	-7.5	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.237	0.263	-11.0	174#	0.00
15 S	2-Fluorobiphenyl	1.518	1.536	-1.2	139	0.01
16	Acenaphthylene	1.864	1.890	-1.4	144	0.00
17	Acenaphthene	1.046	1.052	-0.6	139	0.01
18	Fluorene	1.541	1.556	-1.0	141	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
20	4-Bromophenyl-phenylether	0.231	0.225	2.6	143	0.00
21	Hexachlorobenzene	0.236	0.233	1.3	143	0.00
22	Pentachlorophenol	0.068	0.063	7.4	203#	-0.01
23	Phenanthrene	0.977	0.955	2.3	141	0.00
24	Anthracene	0.886	0.873	1.5	151#	0.00
25 SURR	Fluoranthene-d10	5.466	5.317	2.7	143	0.00
26	Fluoranthene	1.545	1.508	2.4	142	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
28	Pvrene	1.005	1.062	-5.7	144	0.00
29 S	Terphenyl-d14	0.714	0.760	-6.4	148	0.00
30	Benzo(a)anthracene	1.137	1.172	-3.1	145	0.00
31	Chrysene	1.215	1.226	-0.9	136	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.521	-16.1	172#	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.462	0.3	142	0.00
34 I	Perylene-d12	1.000	1.000	0.0	124	0.00
35	Benzo(b)fluoranthene	1.096	1.129	-3.0	142	0.00
36	Benzo(k)fluoranthene	1.143	1.131	1.0	137	0.00
37 C	Benzo(a)pyrene	1.055	1.054	0.1	140	0.00
38	Dibenzo(a,h)anthracene	1.108	1.088	1.8	141	0.00
39	Benzo(a,h,i)perylene	1.118	1.101	1.5	138	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060319\
Data File : BE099790.D
Acq On : 3 Jun 2019 19:17
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 04 03:55:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			