

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060419\
 Data File : BE099792.D
 Acq On : 4 Jun 2019 9:04
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 04 11:53:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.592	0.531	10.3	63	-0.02
3	n-Nitrosodimethylamine	0.327	0.261	20.2	61	-0.02
4 S	2-Fluorophenol	1.011	0.907	10.3	69	0.00
5 S	Phenol-d6	1.338	1.154	13.8	70	0.00
6	bis(2-Chloroethyl)ether	1.190	1.079	9.3	70	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	70	0.00
8 S	Nitrobenzene-d5	0.346	0.293	15.3	69	-0.01
9	Naphthalene	4.278	3.836	10.3	71	-0.01
10	Hexachlorobutadiene	0.315	0.287	8.9	72	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.616	10.7	71	-0.01
12	2-Methylnaphthalene	0.697	0.593	14.9	69	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	-0.01
14 S	2,4,6-Tribromophenol	0.237	0.192	19.0	75	-0.01
15 S	2-Fluorobiphenyl	1.518	1.347	11.3	71	0.00
16	Acenaphthylene	1.864	1.742	6.5	78	-0.01
17	Acenaphthene	1.046	0.917	12.3	71	0.00
18	Fluorene	1.541	1.373	10.9	73	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
20	4-Bromophenyl-phenylether	0.231	0.184	20.3	72	-0.01
21	Hexachlorobenzene	0.236	0.213	9.7	80	-0.01
22	Pentachlorophenol	0.068	0.033	51.5#	67	0.01
23	Phenanthrene	0.977	0.852	12.8	78	0.00
24	Anthracene	0.886	0.734	17.2	78	0.00
25 SURR	Fluoranthene-d10	5.466	5.232	4.3	87	0.00
26	Fluoranthene	1.545	1.504	2.7	88	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
28	Pvrene	1.005	0.932	7.3	90	-0.01
29 S	Terphenyl-d14	0.714	0.627	12.2	87	-0.01
30	Benzo(a)anthracene	1.137	1.057	7.0	93	-0.01
31	Chrysene	1.215	1.197	1.5	95	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.185	9.5	96	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.430	2.5	99	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	96	-0.02
35	Benzo(b)fluoranthene	1.096	1.009	7.9	98	-0.01
36	Benzo(k)fluoranthene	1.143	1.030	9.9	96	-0.02
37 C	Benzo(a)pyrene	1.055	0.955	9.5	98	-0.01
38	Dibenzo(a,h)anthracene	1.108	0.975	12.0	97	-0.02
39	Benzo(a,h,i)perylene	1.118	1.002	10.4	97	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060419\
Data File : BE099792.D
Acq On : 4 Jun 2019 9:04
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 04 11:53:45 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			