

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060419\
 Data File : BE099795.D
 Acq On : 4 Jun 2019 19:14
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 05 03:50:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.592	0.530	10.5	73	-0.01
3	n-Nitrosodimethylamine	0.327	0.276	15.6	75	-0.01
4 S	2-Fluorophenol	1.011	0.887	12.3	79	0.01
5 S	Phenol-d6	1.338	1.163	13.1	82	0.01
6	bis(2-Chloroethyl)ether	1.190	1.032	13.3	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.01
8 S	Nitrobenzene-d5	0.346	0.303	12.4	78	0.00
9	Naphthalene	4.278	3.862	9.7	77	0.01
10	Hexachlorobutadiene	0.315	0.299	5.1	81	0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.601	12.9	75	0.00
12	2-Methylnaphthalene	0.697	0.629	9.8	79	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.237	0.210	11.4	87	0.00
15 S	2-Fluorobiphenyl	1.518	1.331	12.3	76	0.01
16	Acenaphthylene	1.864	1.743	6.5	83	0.00
17	Acenaphthene	1.046	0.961	8.1	80	0.01
18	Fluorene	1.541	1.388	9.9	79	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.231	0.188	18.6	81	0.00
21	Hexachlorobenzene	0.236	0.208	11.9	87	0.00
22	Pentachlorophenol	0.068	0.046	32.4#	101	0.00
23	Phenanthrene	0.977	0.882	9.7	89	0.00
24	Anthracene	0.886	0.755	14.8	88	0.00
25 SURR	Fluoranthene-d10	5.466	5.370	1.8	98	0.00
26	Fluoranthene	1.545	1.518	1.7	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.005	0.982	2.3	99	0.00
29 S	Terphenyl-d14	0.714	0.662	7.3	96	0.00
30	Benzo(a)anthracene	1.137	1.094	3.8	101	0.00
31	Chrysene	1.215	1.167	4.0	97	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.269	3.1	107	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.391	5.2	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.096	1.012	7.7	101	0.00
36	Benzo(k)fluoranthene	1.143	1.044	8.7	100	0.00
37 C	Benzo(a)pyrene	1.055	0.958	9.2	101	0.00
38	Dibenzo(a,h)anthracene	1.108	0.954	13.9	98	0.00
39	Benzo(a,h,i)perylene	1.118	1.007	9.9	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060419\
Data File : BE099795.D
Acq On : 4 Jun 2019 19:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 05 03:50:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			