

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
 Data File : BE099827.D
 Acq On : 6 Jun 2019 18:24
 Operator : JU/SJ
 Sample : K3190-06
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BP-TT-AOC22-MW08-20190603

Quant Time: Jun 07 04:37:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	992	0.40	ng	0.00
7) Naphthalene-d8	10.61	136	3208	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	2390	0.40	ng	0.01
19) Phenanthrene-d10	17.23	188	8200	0.40	ng	0.00
27) Chrysene-d12	21.40	240	15332	0.40	ng	-0.01
34) Perylene-d12	23.94	264	18079	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.37	112	380	0.16	ng	0.01
5) Phenol-d6	6.97	99	276	0.09	ng	0.01
8) Nitrobenzene-d5	8.99	82	914	0.32	ng	0.03
11) 2-Methylnaphthalene-d10	12.22	152	1787	0.32	ng	0.01
14) 2,4,6-Tribromophenol	15.98	330	413	0.36	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	3411	0.38	ng	0.01
25) Fluoranthene-d10	19.26	212	51548	0.42	ng	0.00
29) Terphenyl-d14	19.85	244	12658	0.50	ng	0.00
Target Compounds						
32) Bis(2-ethylhexyl)phthalate	21.31	149	1960	0.040	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
Data File : BE099827.D
Acq On : 6 Jun 2019 18:24
Operator : JU/SJ
Sample : K3190-06
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BP-TT-AOC22-MW08-20190603

Quant Time: Jun 07 04:37:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

