

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
 Data File : BE099829.D
 Acq On : 6 Jun 2019 19:36
 Operator : JU/SJ
 Sample : K3190-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BP-TT-AOC22-MW04-20190603

Quant Time: Jun 07 04:46:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	1248	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	4769	0.40	ng	-0.01
13) Acenaphthene-d10	14.47	164	6304	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	15986	0.40	ng	-0.01
27) Chrysene-d12	21.40	240	20251	0.40	ng	-0.01
34) Perylene-d12	23.94	264	22674	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	701	0.23	ng	0.00
5) Phenol-d6	6.96	99	516	0.13	ng	0.00
8) Nitrobenzene-d5	8.96	82	1546	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.21	152	3021	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.96	330	1741	0.49	ng	-0.03
15) 2-Fluorobiphenyl	13.09	172	6553	0.28	ng	-0.02
25) Fluoranthene-d10	19.25	212	82955	0.35	ng	0.00
29) Terphenyl-d14	19.85	244	23589	0.71	ng	0.00
Target Compounds						
9) Naphthalene	10.65	128	1048	0.021	ng	Qvalue # 74
22) Pentachlorophenol	16.87	266	308	0.376	ng	90
32) Bis(2-ethylhexyl)phthalate	21.30	149	11736	0.183	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
Data File : BE099829.D
Acq On : 6 Jun 2019 19:36
Operator : JU/SJ
Sample : K3190-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BP-TT-AOC22-MW04-20190603

Quant Time: Jun 07 04:46:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

