

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
 Data File : BE099814.D
 Acq On : 6 Jun 2019 8:27
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 07 03:50:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.580	0.523	9.8	114	0.00
3	n-Nitrosodimethylamine	0.308	0.257	16.6	110	0.01
4 S	2-Fluorophenol	0.974	0.915	6.1	118	0.00
5 S	Phenol-d6	1.252	1.282	-2.4	124	0.00
6	bis(2-Chloroethyl)ether	1.097	1.044	4.8	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.351	0.324	7.7	115	0.01
9	Naphthalene	4.271	3.998	6.4	114	0.00
10	Hexachlorobutadiene	0.325	0.306	5.8	114	0.01
11 SURR	2-Methylnaphthalene-d10	0.687	0.605	11.9	108	0.01
12	2-Methylnaphthalene	0.681	0.604	11.3	111	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.01
14 S	2,4,6-Tribromophenol	0.233	0.196	15.9	106	0.00
15 S	2-Fluorobiphenyl	1.487	1.341	9.8	113	0.00
16	Acenaphthylene	1.930	1.713	11.2	108	0.00
17	Acenaphthene	1.056	0.937	11.3	108	0.00
18	Fluorene	1.556	1.386	10.9	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.223	0.196	12.1	109	0.01
21	Hexachlorobenzene	0.245	0.211	13.9	102	0.00
22	Pentachlorophenol	0.045	0.039	13.3	127	0.00
23	Phenanthrene	0.955	0.860	9.9	103	0.00
24	Anthracene	0.863	0.726	15.9	102	0.00
25 SURR	Fluoranthene-d10	5.945	5.165	13.1	101	0.00
26	Fluoranthene	1.673	1.523	9.0	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pyrene	0.966	0.972	-0.6	105	0.00
29 S	Terphenyl-d14	0.656	0.660	-0.6	106	0.00
30	Benzo(a)anthracene	1.105	1.112	-0.6	108	0.00
31	Chrysene	1.221	1.166	4.5	99	0.00
32	Bis(2-ethylhexyl)phthalate	1.269	1.288	-1.5	108	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.442	1.4	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.107	1.026	7.3	106	0.00
36	Benzo(k)fluoranthene	1.176	1.043	11.3	101	0.00
37 C	Benzo(a)pyrene	1.074	0.981	8.7	104	0.00
38	Dibenzo(a,h)anthracene	1.092	0.966	11.5	103	0.01
39	Benzo(a,h,i)perylene	1.147	1.014	11.6	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
Data File : BE099814.D
Acq On : 6 Jun 2019 8:27
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 07 03:50:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		