

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
 Data File : BE099824.D
 Acq On : 6 Jun 2019 16:35
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 07 02:06:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.580	0.531	8.4	114	0.00
3	n-Nitrosodimethylamine	0.308	0.264	14.3	112	0.00
4 S	2-Fluorophenol	0.974	1.001	-2.8	127	0.00
5 S	Phenol-d6	1.252	1.272	-1.6	121	0.00
6	bis(2-Chloroethyl)ether	1.097	1.020	7.0	116	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.351	0.327	6.8	115	0.00
9	Naphthalene	4.271	3.916	8.3	111	0.00
10	Hexachlorobutadiene	0.325	0.290	10.8	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.687	0.585	14.8	104	0.00
12	2-Methylnaphthalene	0.681	0.605	11.2	110	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.233	0.226	3.0	118	-0.01
15 S	2-Fluorobiphenyl	1.487	1.386	6.8	112	0.00
16	Acenaphthylene	1.930	1.727	10.5	105	0.00
17	Acenaphthene	1.056	0.960	9.1	106	0.00
18	Fluorene	1.556	1.424	8.5	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.223	0.184	17.5	105	0.00
21	Hexachlorobenzene	0.245	0.209	14.7	104	0.00
22	Pentachlorophenol	0.045	0.060	-33.3#	199#	-0.03
23	Phenanthrene	0.955	0.856	10.4	106	0.00
24	Anthracene	0.863	0.756	12.4	109	0.00
25 SURR	Fluoranthene-d10	5.945	5.167	13.1	104	0.00
26	Fluoranthene	1.673	1.506	10.0	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
28	Pvrene	0.966	0.976	-1.0	108	0.00
29 S	Terphenyl-d14	0.656	0.677	-3.2	111	0.00
30	Benzo(a)anthracene	1.105	1.088	1.5	108	0.00
31	Chrysene	1.221	1.213	0.7	105	-0.01
32	Bis(2-ethylhexyl)phthalate	1.269	1.407	-10.9	120	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.430	2.3	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.107	1.024	7.5	106	0.00
36	Benzo(k)fluoranthene	1.176	1.083	7.9	105	0.00
37 C	Benzo(a)pyrene	1.074	1.000	6.9	106	0.00
38	Dibenzo(a,h)anthracene	1.092	0.996	8.8	107	0.00
39	Benzo(a,h,i)perylene	1.147	1.026	10.5	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
Data File : BE099824.D
Acq On : 6 Jun 2019 16:35
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 07 02:06:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			