

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
 Data File : BE099843.D
 Acq On : 7 Jun 2019 4:03
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 07 05:45:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.580	0.405	30.2#	84	0.01
3	n-Nitrosodimethylamine	0.308	0.272	11.7	111	0.01
4 S	2-Fluorophenol	0.974	0.998	-2.5	123	0.01
5 S	Phenol-d6	1.252	1.503	-20.0	138	0.01
6	bis(2-Chloroethyl)ether	1.097	1.035	5.7	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.351	0.356	-1.4	126	0.00
9	Naphthalene	4.271	4.184	2.0	119	0.01
10	Hexachlorobutadiene	0.325	0.302	7.1	113	0.01
11 SURR	2-Methylnaphthalene-d10	0.687	0.658	4.2	117	0.01
12	2-Methylnaphthalene	0.681	0.695	-2.1	127	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.01
14 S	2,4,6-Tribromophenol	0.233	0.315	-35.2#	177#	0.00
15 S	2-Fluorobiphenyl	1.487	1.347	9.4	117	0.00
16	Acenaphthylene	1.930	1.931	-0.1	126	0.00
17	Acenaphthene	1.056	1.110	-5.1	132	0.00
18	Fluorene	1.556	1.703	-9.4	133	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.223	0.222	0.4	110	0.01
21	Hexachlorobenzene	0.245	0.214	12.7	92	0.01
22	Pentachlorophenol	0.045	0.106	-135.6#	305#	-0.03
23	Phenanthrene	0.955	1.110	-16.2	119	0.00
24	Anthracene	0.863	1.010	-17.0	126	0.00
25 SURR	Fluoranthene-d10	5.945	4.527	23.9	79	0.00
26	Fluoranthene	1.673	1.638	2.1	100	0.01
27 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
28	Pvrene	0.966	1.271	-31.6#	97	0.01
29 S	Terphenyl-d14	0.656	0.782	-19.2	88	0.00
30	Benzo(a)anthracene	1.105	1.403	-27.0#	96	0.01
31	Chrysene	1.221	1.226	-0.4	73	0.01
32	Bis(2-ethylhexyl)phthalate	1.269	2.092	-64.9#	123	0.01
33	Indeno(1,2,3-cd)pyrene	1.463	1.007	31.2#	51	0.05
34 I	Perylene-d12	1.000	1.000	0.0	70	0.03
35	Benzo(b)fluoranthene	1.107	1.125	-1.6	79	0.02
36	Benzo(k)fluoranthene	1.176	1.047	11.0	69	0.02
37 C	Benzo(a)pyrene	1.074	1.077	-0.3	78	0.02
38	Dibenzo(a,h)anthracene	1.092	0.770	29.5#	56	0.04
39	Benzo(a,h,i)perylene	1.147	0.764	33.4#	53	0.06

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060619\
Data File : BE099843.D
Acq On : 7 Jun 2019 4:03
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 07 05:45:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			