

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060719\
 Data File : BE099873.D
 Acq On : 8 Jun 2019 5:58
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 08 16:52:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 15:25:21 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	-0.01
2	1,4-Dioxane	0.596	0.532	10.7	129	0.00
3	n-Nitrosodimethylamine	0.382	0.278	27.2#	94	0.00
4 S	2-Fluorophenol	1.339	1.153	13.9	110	0.00
5 S	Phenol-d6	1.650	1.392	15.6	102	0.00
6	bis(2-Chloroethyl)ether	1.213	1.162	4.2	126	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	127	-0.01
8 S	Nitrobenzene-d5	0.403	0.371	7.9	122	0.00
9	Naphthalene	4.347	4.354	-0.2	129	0.00
10	Hexachlorobutadiene	0.320	0.322	-0.6	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.724	0.697	3.7	123	0.00
12	2-Methylnaphthalene	0.739	0.688	6.9	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.396	0.286	27.8#	78	-0.01
15 S	2-Fluorobiphenyl	1.446	1.501	-3.8	124	0.00
16	Acenaphthylene	1.981	1.951	1.5	116	0.00
17	Acenaphthene	1.086	1.083	0.3	118	-0.02
18	Fluorene	1.661	1.589	4.3	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	-0.01
20	4-Bromophenyl-phenylether	0.229	0.219	4.4	107	0.00
21	Hexachlorobenzene	0.224	0.226	-0.9	109	0.00
22	Pentachlorophenol	0.069	0.066	4.3	526#	-0.03
23	Phenanthrene	0.992	0.951	4.1	102	-0.01
24	Anthracene	0.951	0.825	13.2	93	-0.01
25 SURR	Fluoranthene-d10	6.199	5.601	9.6	95	0.00
26	Fluoranthene	1.751	1.615	7.8	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.059	1.086	-2.5	92	0.00
29 S	Terphenyl-d14	0.746	0.728	2.4	90	0.00
30	Benzo(a)anthracene	1.271	1.275	-0.3	95	0.00
31	Chrysene	1.248	1.376	-10.3	103	-0.01
32	Bis(2-ethylhexyl)phthalate	2.073	1.725	16.8	80	-0.01
33	Indeno(1,2,3-cd)pyrene	1.520	1.611	-6.0	101	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
35	Benzo(b)fluoranthene	1.111	1.076	3.2	100	-0.01
36	Benzo(k)fluoranthene	1.111	1.134	-2.1	102	0.00
37 C	Benzo(a)pyrene	1.078	1.065	1.2	100	-0.02
38	Dibenzo(a,h)anthracene	1.102	1.089	1.2	101	-0.01
39	Benzo(a,h,i)perylene	1.120	1.135	-1.3	103	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060719\
Data File : BE099873.D
Acq On : 8 Jun 2019 5:58
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 08 16:52:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 15:25:21 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			