

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE060719\
 Data File : BE099882.D
 Acq On : 8 Jun 2019 12:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 08 16:54:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 15:25:21 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	149	0.00
2	1,4-Dioxane	0.400	0.363	9.3	153	0.00
3	n-Nitrosodimethylamine	0.400	0.359	10.3	143	0.00
4 S	2-Fluorophenol	0.400	0.336	16.0	132	0.00
5 S	Phenol-d6	0.400	0.346	13.5	128	0.00
6	bis(2-Chloroethyl)ether	0.400	0.369	7.8	149	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	152	-0.01
8 S	Nitrobenzene-d5	0.400	0.389	2.8	154	0.00
9	Naphthalene	0.400	0.395	1.3	152	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	157	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	146	0.00
12	2-Methylnaphthalene	0.400	0.368	8.0	140	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.400	0.277	30.8#	85	-0.01
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	142	0.00
16	Acenaphthylene	0.400	0.409	-2.2	136	0.00
17	Acenaphthene	0.400	0.400	0.0	134	-0.02
18	Fluorene	0.400	0.383	4.3	124	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	-0.01
20	4-Bromophenyl-phenylether	0.400	0.403	-0.8	123	0.00
21	Hexachlorobenzene	0.400	0.432	-8.0	128	0.00
22	Pentachlorophenol	0.400	0.683	-70.8#	830	-0.04
23	Phenanthrene	0.400	0.392	2.0	114	-0.01
24	Anthracene	0.400	0.369	7.8	108	-0.01
25 SURR	Fluoranthene-d10	0.400	0.365	8.8	104	0.00
26	Fluoranthene	0.400	0.372	7.0	106	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
28	Pvrene	0.400	0.418	-4.5	102	0.00
29 S	Terphenyl-d14	0.400	0.404	-1.0	101	0.00
30	Benzo(a)anthracene	0.400	0.412	-3.0	106	-0.01
31	Chrysene	0.400	0.429	-7.2	108	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.331	17.3	86	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	109	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	108	0.00
35	Benzo(b)fluoranthene	0.400	0.400	0.0	110	-0.01
36	Benzo(k)fluoranthene	0.400	0.411	-2.7	110	-0.01
37 C	Benzo(a)pyrene	0.400	0.403	-0.8	110	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.400	0.0	109	-0.02
39	Benzo(a,h,i)perylene	0.400	0.406	-1.5	110	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE060719\
Data File : BE099882.D
Acq On : 8 Jun 2019 12:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 08 16:54:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 15:25:21 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				