

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE060719\
 Data File : BE099882.D
 Acq On : 8 Jun 2019 12:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 08 16:54:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 07 15:25:21 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	149	0.00
2	1,4-Dioxane	0.596	0.514	13.8	153#	0.00
3	n-Nitrosodimethylamine	0.382	0.342	10.5	143	0.00
4 S	2-Fluorophenol	1.339	1.126	15.9	132	0.00
5 S	Phenol-d6	1.650	1.426	13.6	128	0.00
6	bis(2-Chloroethyl)ether	1.213	1.119	7.7	149	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	152#	-0.01
8 S	Nitrobenzene-d5	0.403	0.392	2.7	154#	0.00
9	Naphthalene	4.347	4.294	1.2	152#	0.00
10	Hexachlorobutadiene	0.320	0.329	-2.8	157#	0.00
11 SURR	2-Methylnaphthalene-d10	0.724	0.692	4.4	146	0.00
12	2-Methylnaphthalene	0.739	0.679	8.1	140	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	130	0.00
14 S	2,4,6-Tribromophenol	0.396	0.274	30.8#	85	-0.01
15 S	2-Fluorobiphenyl	1.446	1.512	-4.6	142	0.00
16	Acenaphthylene	1.981	2.025	-2.2	136	0.00
17	Acenaphthene	1.086	1.086	0.0	134	-0.02
18	Fluorene	1.661	1.590	4.3	124	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	-0.01
20	4-Bromophenyl-phenylether	0.229	0.231	-0.9	123	0.00
21	Hexachlorobenzene	0.224	0.241	-7.6	128	0.00
22	Pentachlorophenol	0.069	0.095	-37.7#	830#	-0.04
23	Phenanthrene	0.992	0.971	2.1	114	-0.01
24	Anthracene	0.951	0.876	7.9	108	-0.01
25 SURR	Fluoranthene-d10	6.199	5.654	8.8	104	0.00
26	Fluoranthene	1.751	1.627	7.1	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	-0.01
28	Pvrene	1.059	1.106	-4.4	102	0.00
29 S	Terphenyl-d14	0.746	0.754	-1.1	101	0.00
30	Benzo(a)anthracene	1.271	1.308	-2.9	106	-0.01
31	Chrysene	1.248	1.338	-7.2	108	-0.01
32	Bis(2-ethylhexyl)phthalate	2.073	1.713	17.4	86	-0.01
33	Indeno(1,2,3-cd)pyrene	1.520	1.607	-5.7	109	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.111	1.111	0.0	110	-0.01
36	Benzo(k)fluoranthene	1.111	1.142	-2.8	110	-0.01
37 C	Benzo(a)pyrene	1.078	1.085	-0.6	110	-0.01
38	Dibenzo(a,h)anthracene	1.102	1.102	0.0	109	-0.02
39	Benzo(a,h,i)perylene	1.120	1.136	-1.4	110	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE060719\
Data File : BE099882.D
Acq On : 8 Jun 2019 12:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 08 16:54:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 07 15:25:21 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			