

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
 Data File : BE099908.D
 Acq On : 11 Jun 2019 1:57
 Operator : JU/SJ
 Sample : K3257-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE103D3-20190605

Quant Time: Jun 11 07:43:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	706	0.40	ng	0.00
7) Naphthalene-d8	10.61	136	2283	0.40	ng	0.01
13) Acenaphthene-d10	14.47	164	1929	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	6080	0.40	ng	0.00
27) Chrysene-d12	21.40	240	9101	0.40	ng	0.00
34) Perylene-d12	23.93	264	10314	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	267	0.14	ng	0.00
5) Phenol-d6	6.99	99	155	0.06	ng	0.02
8) Nitrobenzene-d5	8.99	82	800	0.37	ng	0.03
11) 2-Methylnaphthalene-d10	12.21	152	1527	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	395	0.28	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	3060	0.42	ng	0.00
25) Fluoranthene-d10	19.25	212	36589	0.41	ng	0.00
29) Terphenyl-d14	19.84	244	9563	0.58	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.27	88	457	0.437	ng	Qvalue # 81
32) Bis(2-ethylhexyl)phthalate	21.29	149	6232	0.160	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
Data File : BE099908.D
Acq On : 11 Jun 2019 1:57
Operator : JU/SJ
Sample : K3257-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE103D3-20190605

Quant Time: Jun 11 07:43:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

