

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
 Data File : BE099892.D
 Acq On : 10 Jun 2019 16:16
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE061019

Quant Time: Jun 10 18:12:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 18:07:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.02
2	1,4-Dioxane	0.400	0.414	-3.5	128	0.00
3	n-Nitrosodimethylamine	0.400	0.370	7.5	107	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	109	0.02
5 S	Phenol-d6	0.400	0.377	5.8	118	0.02
6	bis(2-Chloroethyl)ether	0.400	0.426	-6.5	130	0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.01
8 S	Nitrobenzene-d5	0.400	0.372	7.0	111	0.01
9	Naphthalene	0.400	0.400	0.0	114	0.01
10	Hexachlorobutadiene	0.400	0.408	-2.0	112	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.397	0.8	114	0.01
12	2-Methylnaphthalene	0.400	0.396	1.0	113	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	0.01
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	105	0.01
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	114	0.01
16	Acenaphthylene	0.400	0.389	2.8	112	0.00
17	Acenaphthene	0.400	0.387	3.3	112	0.01
18	Fluorene	0.400	0.389	2.8	114	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.01
20	4-Bromophenyl-phenylether	0.400	0.395	1.3	119	0.01
21	Hexachlorobenzene	0.400	0.396	1.0	116	0.01
22	Pentachlorophenol	0.400	0.441	-10.2	68	0.03
23	Phenanthrene	0.400	0.388	3.0	115	0.01
24	Anthracene	0.400	0.372	7.0	109	0.01
25 SURR	Fluoranthene-d10	0.400	0.369	7.8	105	0.00
26	Fluoranthene	0.400	0.369	7.8	104	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	99	0.00
28	Pvrene	0.400	0.451	-12.7	104	0.00
29 S	Terphenyl-d14	0.400	0.429	-7.2	101	0.00
30	Benzo(a)anthracene	0.400	0.402	-0.5	96	0.01
31	Chrysene	0.400	0.446	-11.5	101	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.406	-1.5	99	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	97	0.00
34 I	Perylene-d12	0.400	0.400	0.0	97	0.01
35	Benzo(b)fluoranthene	0.400	0.382	4.5	97	0.00
36	Benzo(k)fluoranthene	0.400	0.449	-12.2	109	0.00
37 C	Benzo(a)pyrene	0.400	0.390	2.5	97	0.01
38	Dibenzo(a,h)anthracene	0.400	0.380	5.0	96	0.02
39	Benzo(a,h,i)perylene	0.400	0.398	0.5	99	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
Data File : BE099892.D
Acq On : 10 Jun 2019 16:16
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE061019

Quant Time: Jun 10 18:12:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 18:07:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			