

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
 Data File : BE099892.D
 Acq On : 10 Jun 2019 16:16
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE061019

Quant Time: Jun 10 18:12:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 18:07:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.02
2	1,4-Dioxane	0.592	0.613	-3.5	128	0.00
3	n-Nitrosodimethylamine	0.341	0.316	7.3	107	0.00
4 S	2-Fluorophenol	1.080	1.005	6.9	109	0.02
5 S	Phenol-d6	1.391	1.312	5.7	118	0.02
6	bis(2-Chloroethyl)ether	1.136	1.210	-6.5	130	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.01
8 S	Nitrobenzene-d5	0.379	0.352	7.1	111	0.01
9	Naphthalene	4.318	4.318	0.0	114	0.01
10	Hexachlorobutadiene	0.329	0.336	-2.1	112	0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.693	0.7	114	0.01
12	2-Methylnaphthalene	0.702	0.695	1.0	113	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.01
14 S	2,4,6-Tribromophenol	0.296	0.251	15.2	105	0.01
15 S	2-Fluorobiphenyl	1.517	1.471	3.0	114	0.01
16	Acenaphthylene	1.990	1.937	2.7	112	0.00
17	Acenaphthene	1.084	1.048	3.3	112	0.01
18	Fluorene	1.627	1.581	2.8	114	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.01
20	4-Bromophenyl-phenylether	0.227	0.224	1.3	119	0.01
21	Hexachlorobenzene	0.232	0.229	1.3	116	0.01
22	Pentachlorophenol	0.073	0.039	46.6#	68	0.03
23	Phenanthrene	0.971	0.940	3.2	115	0.01
24	Anthracene	0.890	0.827	7.1	109	0.01
25 SURR	Fluoranthene-d10	5.828	5.371	7.8	105	0.00
26	Fluoranthene	1.650	1.522	7.8	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.051	1.185	-12.7	104	0.00
29 S	Terphenyl-d14	0.720	0.772	-7.2	101	0.00
30	Benzo(a)anthracene	1.195	1.202	-0.6	96	0.01
31	Chrysene	1.236	1.377	-11.4	101	0.01
32	Bis(2-ethylhexyl)phthalate	1.716	1.743	-1.6	99	0.01
33	Indeno(1,2,3-cd)pyrene	1.466	1.524	-4.0	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.01
35	Benzo(b)fluoranthene	1.121	1.069	4.6	97	0.00
36	Benzo(k)fluoranthene	1.172	1.317	-12.4	109	0.00
37 C	Benzo(a)pyrene	1.104	1.077	2.4	97	0.01
38	Dibenzo(a,h)anthracene	1.112	1.057	4.9	96	0.02
39	Benzo(a,h,i)perylene	1.149	1.142	0.6	99	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
Data File : BE099892.D
Acq On : 10 Jun 2019 16:16
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE061019

Quant Time: Jun 10 18:12:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 18:07:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			