

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
 Data File : BE099911.D
 Acq On : 11 Jun 2019 3:44
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 11 07:48:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.592	0.599	-1.2	134	0.00
3	n-Nitrosodimethylamine	0.341	0.201	41.1#	73	0.00
4 S	2-Fluorophenol	1.080	0.915	15.3	106	0.00
5 S	Phenol-d6	1.391	1.281	7.9	123	0.00
6	bis(2-Chloroethyl)ether	1.136	1.137	-0.1	131	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
8 S	Nitrobenzene-d5	0.379	0.368	2.9	124	0.00
9	Naphthalene	4.318	4.365	-1.1	123	0.00
10	Hexachlorobutadiene	0.329	0.347	-5.5	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.698	0.683	2.1	120	0.00
12	2-Methylnaphthalene	0.702	0.682	2.8	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.296	0.231	22.0	98	0.00
15 S	2-Fluorobiphenyl	1.517	1.511	0.4	119	0.00
16	Acenaphthylene	1.990	1.922	3.4	113	-0.01
17	Acenaphthene	1.084	1.044	3.7	113	0.00
18	Fluorene	1.627	1.525	6.3	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.227	0.220	3.1	115	0.00
21	Hexachlorobenzene	0.232	0.223	3.9	111	0.00
22	Pentachlorophenol	0.073	0.060	17.8	103	0.00
23	Phenanthrene	0.971	0.917	5.6	110	0.00
24	Anthracene	0.890	0.769	13.6	100	0.00
25 SURR	Fluoranthene-d10	5.828	5.211	10.6	100	0.00
26	Fluoranthene	1.650	1.481	10.2	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.051	1.099	-4.6	101	0.00
29 S	Terphenyl-d14	0.720	0.731	-1.5	100	0.00
30	Benzo(a)anthracene	1.195	1.239	-3.7	103	0.00
31	Chrysene	1.236	1.323	-7.0	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.588	7.5	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.563	-6.6	104	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.121	1.112	0.8	106	0.00
36	Benzo(k)fluoranthene	1.172	1.157	1.3	101	0.00
37 C	Benzo(a)pyrene	1.104	1.084	1.8	103	0.00
38	Dibenzo(a,h)anthracene	1.112	1.065	4.2	103	0.00
39	Benzo(a,h,i)perylene	1.149	1.141	0.7	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061019\
Data File : BE099911.D
Acq On : 11 Jun 2019 3:44
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 11 07:48:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		