

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
 Data File : BE099913.D
 Acq On : 11 Jun 2019 8:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 03:53:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.592	0.552	6.8	95	0.00
3	n-Nitrosodimethylamine	0.341	0.279	18.2	78	0.02
4 S	2-Fluorophenol	1.080	0.880	18.5	78	0.01
5 S	Phenol-d6	1.391	1.246	10.4	92	0.00
6	bis(2-Chloroethyl)ether	1.136	1.038	8.6	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
8 S	Nitrobenzene-d5	0.379	0.343	9.5	90	0.01
9	Naphthalene	4.318	4.416	-2.3	97	0.00
10	Hexachlorobutadiene	0.329	0.350	-6.4	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.698	0.722	-3.4	99	0.00
12	2-Methylnaphthalene	0.702	0.706	-0.6	95	0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.296	0.211	28.7#	79	0.01
15 S	2-Fluorobiphenyl	1.517	1.411	7.0	98	0.00
16	Acenaphthylene	1.990	1.944	2.3	100	0.00
17	Acenaphthene	1.084	1.012	6.6	96	0.00
18	Fluorene	1.627	1.577	3.1	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.227	0.213	6.2	106	0.00
21	Hexachlorobenzene	0.232	0.222	4.3	105	0.00
22	Pentachlorophenol	0.073	0.034	53.4#	56	0.04
23	Phenanthrene	0.971	0.929	4.3	106	0.00
24	Anthracene	0.890	0.777	12.7	96	0.00
25 SURR	Fluoranthene-d10	5.828	5.401	7.3	98	0.00
26	Fluoranthene	1.650	1.545	6.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.051	1.163	-10.7	99	0.00
29 S	Terphenyl-d14	0.720	0.752	-4.4	95	0.00
30	Benzo(a)anthracene	1.195	1.224	-2.4	94	0.01
31	Chrysene	1.236	1.350	-9.2	96	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.476	14.0	81	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.525	-4.0	94	0.00
34 I	Perylene-d12	1.000	1.000	0.0	93	0.00
35	Benzo(b)fluoranthene	1.121	1.103	1.6	96	0.00
36	Benzo(k)fluoranthene	1.172	1.201	-2.5	96	0.00
37 C	Benzo(a)pyrene	1.104	1.098	0.5	96	0.00
38	Dibenzo(a,h)anthracene	1.112	1.049	5.7	92	0.00
39	Benzo(a,h,i)perylene	1.149	1.156	-0.6	96	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
Data File : BE099913.D
Acq On : 11 Jun 2019 8:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 03:53:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			