

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
 Data File : BE099925.D
 Acq On : 11 Jun 2019 15:43
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 04:17:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	-0.01
2	1,4-Dioxane	0.592	0.619	-4.6	111	0.00
3	n-Nitrosodimethylamine	0.341	0.240	29.6#	69	0.00
4 S	2-Fluorophenol	1.080	1.009	6.6	93	0.00
5 S	Phenol-d6	1.391	1.420	-2.1	109	0.00
6	bis(2-Chloroethyl)ether	1.136	1.175	-3.4	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.01
8 S	Nitrobenzene-d5	0.379	0.333	12.1	97	-0.01
9	Naphthalene	4.318	4.509	-4.4	110	-0.03
10	Hexachlorobutadiene	0.329	0.346	-5.2	107	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.721	-3.3	110	-0.03
12	2-Methylnaphthalene	0.702	0.713	-1.6	107	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	-0.03
14 S	2,4,6-Tribromophenol	0.296	0.217	26.7#	94	-0.03
15 S	2-Fluorobiphenyl	1.517	1.425	6.1	115	-0.03
16	Acenaphthylene	1.990	1.928	3.1	116	-0.03
17	Acenaphthene	1.084	1.037	4.3	115	-0.03
18	Fluorene	1.627	1.664	-2.3	125	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.03
20	4-Bromophenyl-phenylether	0.227	0.217	4.4	126	-0.03
21	Hexachlorobenzene	0.232	0.229	1.3	127	-0.04
22	Pentachlorophenol	0.073	0.034	53.4#	64	-0.04
23	Phenanthrene	0.971	0.933	3.9	125	-0.03
24	Anthracene	0.890	0.807	9.3	116	-0.03
25 SURR	Fluoranthene-d10	5.828	5.087	12.7	108	-0.04
26	Fluoranthene	1.650	1.468	11.0	110	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	101	-0.04
28	Pvrene	1.051	1.211	-15.2	109	-0.03
29 S	Terphenyl-d14	0.720	0.784	-8.9	105	-0.03
30	Benzo(a)anthracene	1.195	1.214	-1.6	99	-0.04
31	Chrysene	1.236	1.321	-6.9	99	-0.04
32	Bis(2-ethylhexyl)phthalate	1.716	1.612	6.1	93	-0.04
33	Indeno(1,2,3-cd)pyrene	1.466	1.505	-2.7	98	-0.10
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.07
35	Benzo(b)fluoranthene	1.121	1.068	4.7	100	-0.06
36	Benzo(k)fluoranthene	1.172	1.210	-3.2	103	-0.06
37 C	Benzo(a)pyrene	1.104	1.080	2.2	101	-0.06
38	Dibenzo(a,h)anthracene	1.112	0.992	10.8	94	-0.09
39	Benzo(a,h,i)perylene	1.149	1.135	1.2	101	-0.11

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
Data File : BE099925.D
Acq On : 11 Jun 2019 15:43
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 04:17:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			