

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
 Data File : BE099927.D
 Acq On : 11 Jun 2019 16:57
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 12 16:38:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	-0.01
2	1,4-Dioxane	0.592	0.518	12.5	103	0.00
3	n-Nitrosodimethylamine	0.341	0.306	10.3	99	0.03
4 S	2-Fluorophenol	1.080	0.936	13.3	96	0.00
5 S	Phenol-d6	1.391	1.318	5.2	113	0.00
6	bis(2-Chloroethyl)ether	1.136	1.105	2.7	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	-0.01
8 S	Nitrobenzene-d5	0.379	0.337	11.1	104	0.00
9	Naphthalene	4.318	4.441	-2.8	115	-0.03
10	Hexachlorobutadiene	0.329	0.343	-4.3	113	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.710	-1.7	114	-0.01
12	2-Methylnaphthalene	0.702	0.713	-1.6	113	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	-0.03
14 S	2,4,6-Tribromophenol	0.296	0.257	13.2	116	-0.03
15 S	2-Fluorobiphenyl	1.517	1.434	5.5	120	-0.03
16	Acenaphthylene	1.990	1.936	2.7	121	-0.03
17	Acenaphthene	1.084	1.014	6.5	117	-0.03
18	Fluorene	1.627	1.594	2.0	124	-0.04
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	-0.03
20	4-Bromophenyl-phenylether	0.227	0.218	4.0	128	-0.03
21	Hexachlorobenzene	0.232	0.233	-0.4	130	-0.04
22	Pentachlorophenol	0.073	0.032	56.2#	62	-0.03
23	Phenanthrene	0.971	0.921	5.1	125	-0.03
24	Anthracene	0.890	0.779	12.5	114	-0.03
25 SURR	Fluoranthene-d10	5.828	5.070	13.0	109	-0.04
26	Fluoranthene	1.650	1.459	11.6	110	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	100	-0.04
28	Pvrene	1.051	1.226	-16.7	109	-0.04
29 S	Terphenyl-d14	0.720	0.797	-10.7	105	-0.03
30	Benzo(a)anthracene	1.195	1.231	-3.0	99	-0.04
31	Chrysene	1.236	1.306	-5.7	97	-0.04
32	Bis(2-ethylhexyl)phthalate	1.716	1.499	12.6	86	-0.04
33	Indeno(1,2,3-cd)pyrene	1.466	1.511	-3.1	98	-0.11
34 I	Perylene-d12	1.000	1.000	0.0	98	-0.07
35	Benzo(b)fluoranthene	1.121	1.064	5.1	98	-0.06
36	Benzo(k)fluoranthene	1.172	1.144	2.4	96	-0.06
37 C	Benzo(a)pyrene	1.104	1.096	0.7	101	-0.07
38	Dibenzo(a,h)anthracene	1.112	1.002	9.9	93	-0.10
39	Benzo(a,h,i)perylene	1.149	1.138	1.0	100	-0.11

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
Data File : BE099927.D
Acq On : 11 Jun 2019 16:57
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 16:38:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		