

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
 Data File : BE099944.D
 Acq On : 12 Jun 2019 3:09
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 12 05:29:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	-0.02
2	1,4-Dioxane	0.592	0.531	10.3	97	0.00
3	n-Nitrosodimethylamine	0.341	0.185	45.7#	55	0.02
4 S	2-Fluorophenol	1.080	0.984	8.9	93	0.00
5 S	Phenol-d6	1.391	1.342	3.5	106	0.00
6	bis(2-Chloroethyl)ether	1.136	0.978	13.9	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	-0.01
8 S	Nitrobenzene-d5	0.379	0.326	14.0	99	0.00
9	Naphthalene	4.318	4.353	-0.8	111	-0.03
10	Hexachlorobutadiene	0.329	0.351	-6.7	113	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.708	-1.4	113	-0.01
12	2-Methylnaphthalene	0.702	0.701	0.1	110	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	-0.03
14 S	2,4,6-Tribromophenol	0.296	0.199	32.8#	91	-0.03
15 S	2-Fluorobiphenyl	1.517	1.397	7.9	119	-0.03
16	Acenaphthylene	1.990	1.911	4.0	121	-0.04
17	Acenaphthene	1.084	1.033	4.7	121	-0.03
18	Fluorene	1.627	1.653	-1.6	131	-0.04
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	-0.04
20	4-Bromophenyl-phenylether	0.227	0.219	3.5	133	-0.04
21	Hexachlorobenzene	0.232	0.230	0.9	132	-0.04
22	Pentachlorophenol	0.073	0.026	64.4#	51	0.00
23	Phenanthrene	0.971	0.909	6.4	127	-0.03
24	Anthracene	0.890	0.734	17.5	110	-0.03
25 SURR	Fluoranthene-d10	5.828	4.935	15.3	110	-0.04
26	Fluoranthene	1.650	1.446	12.4	113	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.05
28	Pvrene	1.051	1.243	-18.3	110	-0.04
29 S	Terphenyl-d14	0.720	0.777	-7.9	102	-0.03
30	Benzo(a)anthracene	1.195	1.095	8.4	88	-0.04
31	Chrysene	1.236	1.412	-14.2	105	-0.04
32	Bis(2-ethylhexyl)phthalate	1.716	1.443	15.9	82	-0.04
33	Indeno(1,2,3-cd)pyrene	1.466	1.515	-3.3	97	-0.11
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.07
35	Benzo(b)fluoranthene	1.121	1.104	1.5	102	-0.06
36	Benzo(k)fluoranthene	1.172	1.220	-4.1	103	-0.06
37 C	Benzo(a)pyrene	1.104	1.092	1.1	101	-0.07
38	Dibenzo(a,h)anthracene	1.112	0.957	13.9	89	-0.09
39	Benzo(a,h,i)perylene	1.149	1.123	2.3	99	-0.12

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061119\
Data File : BE099944.D
Acq On : 12 Jun 2019 3:09
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 12 05:29:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			