

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
 Data File : BE099975.D
 Acq On : 13 Jun 2019 2:28
 Operator : JU/SJ
 Sample : K3255-05DL 2X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE122D2-20190604DL

Quant Time: Jun 13 05:54:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	940	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3384	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	2614	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	7601	0.40	ng	0.00
27) Chrysene-d12	21.40	240	10107	0.40	ng	0.00
34) Perylene-d12	23.92	264	11365	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	164	0.06	ng	0.00
5) Phenol-d6	6.98	99	98	0.03	ng	0.01
8) Nitrobenzene-d5	8.99	82	511	0.16	ng	0.03
11) 2-Methylnaphthalene-d10	12.21	152	929	0.16	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	222	0.11	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	1988	0.20	ng	-0.01
25) Fluoranthene-d10	19.25	212	18326	0.17	ng	0.00
29) Terphenyl-d14	19.84	244	4788	0.26	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.27	88	3590	2.580	ng	93
22) Pentachlorophenol	16.88	266	63	0.329	ng	83
32) Bis(2-ethylhexyl)phthalate	21.29	149	2321	0.054	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
Data File : BE099975.D
Acq On : 13 Jun 2019 2:28
Operator : JU/SJ
Sample : K3255-05DL 2X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE122D2-20190604DL

Quant Time: Jun 13 05:54:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

