

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
 Data File : BE099977.D
 Acq On : 13 Jun 2019 3:40
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 04:19:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	143	-0.01
2	1,4-Dioxane	0.592	0.536	9.5	137	0.00
3	n-Nitrosodimethylamine	0.341	0.228	33.1#	95	0.02
4 S	2-Fluorophenol	1.080	1.014	6.1	134	0.00
5 S	Phenol-d6	1.391	1.317	5.3	145	-0.01
6	bis(2-Chloroethyl)ether	1.136	1.122	1.2	148	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	-0.01
8 S	Nitrobenzene-d5	0.379	0.385	-1.6	145	0.00
9	Naphthalene	4.318	4.535	-5.0	142	-0.01
10	Hexachlorobutadiene	0.329	0.375	-14.0	150	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.719	-3.0	141	0.00
12	2-Methylnaphthalene	0.702	0.710	-1.1	137	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.296	0.216	27.0#	104	0.00
15 S	2-Fluorobiphenyl	1.517	1.605	-5.8	143	-0.01
16	Acenaphthylene	1.990	2.046	-2.8	136	-0.01
17	Acenaphthene	1.084	1.083	0.1	133	0.00
18	Fluorene	1.627	1.572	3.4	131	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.227	0.232	-2.2	129	0.00
21	Hexachlorobenzene	0.232	0.246	-6.0	130	0.00
22	Pentachlorophenol	0.073	0.045	38.4#	82	-0.01
23	Phenanthrene	0.971	0.959	1.2	122	0.00
24	Anthracene	0.890	0.839	5.7	116	0.00
25 SURR	Fluoranthene-d10	5.828	5.211	10.6	106	0.00
26	Fluoranthene	1.650	1.514	8.2	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.051	1.248	-18.7	106	0.00
29 S	Terphenyl-d14	0.720	0.810	-12.5	103	0.00
30	Benzo(a)anthracene	1.195	1.342	-12.3	103	0.00
31	Chrysene	1.236	1.403	-13.5	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.600	6.8	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.636	-11.6	101	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	94	0.00
35	Benzo(b)fluoranthene	1.121	1.173	-4.6	103	0.00
36	Benzo(k)fluoranthene	1.172	1.245	-6.2	100	0.00
37 C	Benzo(a)pyrene	1.104	1.153	-4.4	101	0.00
38	Dibenzo(a,h)anthracene	1.112	1.119	-0.6	99	0.00
39	Benzo(a,h,i)perylene	1.149	1.232	-7.2	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
Data File : BE099977.D
Acq On : 13 Jun 2019 3:40
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:19:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			