

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
 Data File : BE099977.D
 Acq On : 13 Jun 2019 3:40
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 04:19:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	143	-0.01
2	1,4-Dioxane	0.400	0.362	9.5	137	0.00
3	n-Nitrosodimethylamine	0.400	0.268	33.0#	95	0.02
4 S	2-Fluorophenol	0.400	0.376	6.0	134	0.00
5 S	Phenol-d6	0.400	0.379	5.3	145	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	148	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	-0.01
8 S	Nitrobenzene-d5	0.400	0.407	-1.7	145	0.00
9	Naphthalene	0.400	0.420	-5.0	142	-0.01
10	Hexachlorobutadiene	0.400	0.456	-14.0	150	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.412	-3.0	141	0.00
12	2-Methylnaphthalene	0.400	0.405	-1.3	137	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.400	0.292	27.0#	104	0.00
15 S	2-Fluorobiphenyl	0.400	0.423	-5.7	143	-0.01
16	Acenaphthylene	0.400	0.411	-2.7	136	-0.01
17	Acenaphthene	0.400	0.400	0.0	133	0.00
18	Fluorene	0.400	0.387	3.3	131	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.400	0.409	-2.2	129	0.00
21	Hexachlorobenzene	0.400	0.424	-6.0	130	0.00
22	Pentachlorophenol	0.400	0.462	-15.5	82	-0.01
23	Phenanthrene	0.400	0.395	1.3	122	0.00
24	Anthracene	0.400	0.377	5.8	116	0.00
25 SURR	Fluoranthene-d10	0.400	0.358	10.5	106	0.00
26	Fluoranthene	0.400	0.367	8.3	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
28	Pvrene	0.400	0.475	-18.7	106	0.00
29 S	Terphenyl-d14	0.400	0.450	-12.5	103	0.00
30	Benzo(a)anthracene	0.400	0.449	-12.2	103	0.00
31	Chrysene	0.400	0.454	-13.5	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.373	6.8	88	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.447	-11.7	101	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	94	0.00
35	Benzo(b)fluoranthene	0.400	0.419	-4.7	103	0.00
36	Benzo(k)fluoranthene	0.400	0.425	-6.2	100	0.00
37 C	Benzo(a)pyrene	0.400	0.418	-4.5	101	0.00
38	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	99	0.00
39	Benzo(a,h,i)perylene	0.400	0.429	-7.2	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061219\
Data File : BE099977.D
Acq On : 13 Jun 2019 3:40
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:19:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				