

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE099987.D
 Acq On : 13 Jun 2019 13:33
 Operator : JU/SJ
 Sample : K3270-06
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 EQUIPMENT-BLANK

Quant Time: Jun 14 03:54:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	669	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	2444	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	2022	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	6777	0.40	ng	0.00
27) Chrysene-d12	21.40	240	9313	0.40	ng	0.00
34) Perylene-d12	23.93	264	10783	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	225	0.12	ng	0.00
5) Phenol-d6	6.96	99	171	0.07	ng	0.00
8) Nitrobenzene-d5	8.98	82	761	0.33	ng	0.01
11) 2-Methylnaphthalene-d10	12.22	152	1176	0.28	ng	0.01
14) 2,4,6-Tribromophenol	16.00	330	26	0.02	ng	0.03
15) 2-Fluorobiphenyl	13.10	172	3691	0.48	ng	0.00
25) Fluoranthene-d10	19.25	212	28316	0.29	ng	0.00
29) Terphenyl-d14	19.84	244	9632	0.57	ng	0.00
Target Compounds						
32) Bis(2-ethylhexyl)phthalate	21.29	149	2021	0.051	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE099987.D
Acq On : 13 Jun 2019 13:33
Operator : JU/SJ
Sample : K3270-06
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
EQUIPMENT-BLANK

Quant Time: Jun 14 03:54:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

