

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE100009.D
 Acq On : 14 Jun 2019 2:48
 Operator : JU/SJ
 Sample : K3279-02
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW312S-20190607

Quant Time: Jun 14 04:44:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	708	0.40	ng	-0.01
7) Naphthalene-d8	10.60	136	2484	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	2044	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	7105	0.40	ng	0.00
27) Chrysene-d12	21.39	240	9732	0.40	ng	-0.01
34) Perylene-d12	23.92	264	11282	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.35	112	258	0.13	ng	0.00
5) Phenol-d6	6.95	99	205	0.08	ng	-0.01
8) Nitrobenzene-d5	8.98	82	754	0.32	ng	0.01
11) 2-Methylnaphthalene-d10	12.21	152	1387	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	448	0.30	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	3025	0.39	ng	-0.01
25) Fluoranthene-d10	19.24	212	33185	0.32	ng	-0.01
29) Terphenyl-d14	19.84	244	10283	0.59	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.27	88	57	0.054	ng	# 8
9) Naphthalene	10.64	128	20600	0.768	ng	98
12) 2-Methylnaphthalene	12.28	142	362	0.083	ng	# 92
17) Acenaphthene	14.53	154	152	0.027	ng	93
18) Fluorene	15.52	166	290	0.035	ng	94
23) Phenanthrene	17.26	178	889	0.052	ng	99
26) Fluoranthene	19.28	202	705	0.024	ng	# 91
28) Pyrene	19.64	202	573	0.022	ng	94
32) Bis(2-ethylhexyl)phthalate	21.29	149	12307	0.295	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE100009.D
Acq On : 14 Jun 2019 2:48
Operator : JU/SJ
Sample : K3279-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW312S-20190607

Quant Time: Jun 14 04:44:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

