

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE099979.D
 Acq On : 13 Jun 2019 8:45
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 14 03:31:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.592	0.555	6.2	95	0.00
3	n-Nitrosodimethylamine	0.341	0.225	34.0#	62	0.03
4 S	2-Fluorophenol	1.080	1.029	4.7	91	0.00
5 S	Phenol-d6	1.391	1.235	11.2	91	0.00
6	bis(2-Chloroethyl)ether	1.136	1.166	-2.6	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.379	0.349	7.9	95	0.01
9	Naphthalene	4.318	4.550	-5.4	104	0.00
10	Hexachlorobutadiene	0.329	0.372	-13.1	108	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.749	-7.3	106	0.00
12	2-Methylnaphthalene	0.702	0.709	-1.0	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.296	0.232	21.6	88	0.00
15 S	2-Fluorobiphenyl	1.517	1.514	0.2	106	0.00
16	Acenaphthylene	1.990	2.081	-4.6	109	-0.01
17	Acenaphthene	1.084	1.117	-3.0	108	0.00
18	Fluorene	1.627	1.716	-5.5	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.227	0.221	2.6	114	0.00
21	Hexachlorobenzene	0.232	0.236	-1.7	116	-0.01
22	Pentachlorophenol	0.073	0.043	41.1#	72	0.00
23	Phenanthrene	0.971	0.965	0.6	115	0.00
24	Anthracene	0.890	0.801	10.0	103	0.00
25 SURR	Fluoranthene-d10	5.828	5.481	6.0	104	0.00
26	Fluoranthene	1.650	1.602	2.9	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.051	1.264	-20.3	105	0.00
29 S	Terphenyl-d14	0.720	0.804	-11.7	99	0.00
30	Benzo(a)anthracene	1.195	1.300	-8.8	98	0.00
31	Chrysene	1.236	1.439	-16.4	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.416	17.5	76	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.635	-11.5	98	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	1.121	1.128	-0.6	97	0.00
36	Benzo(k)fluoranthene	1.172	1.193	-1.8	94	0.00
37 C	Benzo(a)pyrene	1.104	1.130	-2.4	97	0.00
38	Dibenzo(a,h)anthracene	1.112	1.081	2.8	94	0.00
39	Benzo(a,h,i)perylene	1.149	1.221	-6.3	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE099979.D
Acq On : 13 Jun 2019 8:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 14 03:31:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			