

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE099991.D
 Acq On : 13 Jun 2019 15:59
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 14 04:09:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	-0.01
2	1,4-Dioxane	0.592	0.508	14.2	108	0.00
3	n-Nitrosodimethylamine	0.341	0.352	-3.2	121	0.00
4 S	2-Fluorophenol	1.080	1.069	1.0	117	0.00
5 S	Phenol-d6	1.391	1.444	-3.8	132	0.00
6	bis(2-Chloroethyl)ether	1.136	1.128	0.7	124	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	-0.01
8 S	Nitrobenzene-d5	0.379	0.393	-3.7	127	0.00
9	Naphthalene	4.318	4.585	-6.2	123	-0.01
10	Hexachlorobutadiene	0.329	0.381	-15.8	130	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.761	-9.0	128	-0.01
12	2-Methylnaphthalene	0.702	0.752	-7.1	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
14 S	2,4,6-Tribromophenol	0.296	0.295	0.3	131	-0.01
15 S	2-Fluorobiphenyl	1.517	1.557	-2.6	128	-0.01
16	Acenaphthylene	1.990	2.026	-1.8	125	-0.01
17	Acenaphthene	1.084	1.073	1.0	122	0.00
18	Fluorene	1.627	1.688	-3.7	129	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
20	4-Bromophenyl-phenylether	0.227	0.226	0.4	131	-0.01
21	Hexachlorobenzene	0.232	0.247	-6.5	136	-0.01
22	Pentachlorophenol	0.073	0.086	-17.8	163#	-0.03
23	Phenanthrene	0.971	0.964	0.7	129	0.00
24	Anthracene	0.890	0.874	1.8	126	0.00
25 SURR	Fluoranthene-d10	5.828	5.491	5.8	117	0.00
26	Fluoranthene	1.650	1.563	5.3	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pvrene	1.051	1.221	-16.2	117	0.00
29 S	Terphenyl-d14	0.720	0.831	-15.4	118	0.00
30	Benzo(a)anthracene	1.195	1.315	-10.0	114	0.00
31	Chrysene	1.236	1.445	-16.9	115	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.869	-8.9	115	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.701	-16.0	118	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.121	1.141	-1.8	114	0.00
36	Benzo(k)fluoranthene	1.172	1.199	-2.3	109	0.00
37 C	Benzo(a)pyrene	1.104	1.126	-2.0	112	0.00
38	Dibenzo(a,h)anthracene	1.112	1.171	-5.3	118	-0.01
39	Benzo(a,h,i)perylene	1.149	1.192	-3.7	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE099991.D
Acq On : 13 Jun 2019 15:59
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 14 04:09:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			