

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE099993.D
 Acq On : 13 Jun 2019 17:10
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 14 04:12:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.592	0.597	-0.8	95	0.00
3	n-Nitrosodimethylamine	0.341	0.277	18.8	71	0.03
4 S	2-Fluorophenol	1.080	1.084	-0.4	89	0.00
5 S	Phenol-d6	1.391	1.407	-1.2	96	0.00
6	bis(2-Chloroethyl)ether	1.136	1.156	-1.8	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.379	0.379	0.0	95	0.00
9	Naphthalene	4.318	4.664	-8.0	97	-0.01
10	Hexachlorobutadiene	0.329	0.377	-14.6	100	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.769	-10.2	100	0.00
12	2-Methylnaphthalene	0.702	0.756	-7.7	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.296	0.259	12.5	94	0.00
15 S	2-Fluorobiphenyl	1.517	1.516	0.1	103	-0.01
16	Acenaphthylene	1.990	2.027	-1.9	102	-0.01
17	Acenaphthene	1.084	1.098	-1.3	102	0.00
18	Fluorene	1.627	1.721	-5.8	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.227	0.222	2.2	108	0.00
21	Hexachlorobenzene	0.232	0.243	-4.7	112	0.00
22	Pentachlorophenol	0.073	0.061	16.4	97	-0.01
23	Phenanthrene	0.971	0.963	0.8	108	0.00
24	Anthracene	0.890	0.853	4.2	103	0.00
25 SURR	Fluoranthene-d10	5.828	5.441	6.6	97	0.00
26	Fluoranthene	1.650	1.553	5.9	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
28	Pvrene	1.051	1.213	-15.4	96	0.00
29 S	Terphenyl-d14	0.720	0.791	-9.9	93	0.00
30	Benzo(a)anthracene	1.195	1.325	-10.9	95	0.00
31	Chrysene	1.236	1.382	-11.8	92	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.546	9.9	79	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.639	-11.8	94	0.00
34 I	Perylene-d12	1.000	1.000	0.0	87	0.00
35	Benzo(b)fluoranthene	1.121	1.148	-2.4	93	0.00
36	Benzo(k)fluoranthene	1.172	1.206	-2.9	90	0.00
37 C	Benzo(a)pyrene	1.104	1.134	-2.7	92	0.00
38	Dibenzo(a,h)anthracene	1.112	1.132	-1.8	93	0.00
39	Benzo(a,h,i)perylene	1.149	1.209	-5.2	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE099993.D
Acq On : 13 Jun 2019 17:10
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 14 04:12:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			