

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
 Data File : BE100010.D
 Acq On : 14 Jun 2019 3:24
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 14 04:47:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	-0.01
2	1,4-Dioxane	0.592	0.580	2.0	111	0.00
3	n-Nitrosodimethylamine	0.341	0.307	10.0	95	0.01
4 S	2-Fluorophenol	1.080	1.144	-5.9	112	0.00
5 S	Phenol-d6	1.391	1.569	-12.8	128	-0.01
6	bis(2-Chloroethyl)ether	1.136	1.169	-2.9	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	-0.01
8 S	Nitrobenzene-d5	0.379	0.376	0.8	117	0.00
9	Naphthalene	4.318	4.511	-4.5	117	-0.01
10	Hexachlorobutadiene	0.329	0.366	-11.2	121	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.745	-6.7	121	0.00
12	2-Methylnaphthalene	0.702	0.747	-6.4	120	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.296	0.250	15.5	113	0.00
15 S	2-Fluorobiphenyl	1.517	1.522	-0.3	127	-0.01
16	Acenaphthylene	1.990	2.011	-1.1	126	-0.01
17	Acenaphthene	1.084	1.112	-2.6	128	0.00
18	Fluorene	1.627	1.700	-4.5	132	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.227	0.218	4.0	132	-0.01
21	Hexachlorobenzene	0.232	0.236	-1.7	136	-0.01
22	Pentachlorophenol	0.073	0.064	12.3	126	-0.03
23	Phenanthrene	0.971	0.952	2.0	132	0.00
24	Anthracene	0.890	0.859	3.5	129	0.00
25 SURR	Fluoranthene-d10	5.828	5.383	7.6	119	0.00
26	Fluoranthene	1.650	1.563	5.3	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	110	-0.01
28	Pvrene	1.051	1.260	-19.9	124	0.00
29 S	Terphenyl-d14	0.720	0.803	-11.5	117	0.00
30	Benzo(a)anthracene	1.195	1.281	-7.2	114	0.00
31	Chrysene	1.236	1.460	-18.1	120	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.735	-1.1	109	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.593	-8.7	113	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.121	1.191	-6.2	119	0.00
36	Benzo(k)fluoranthene	1.172	1.258	-7.3	115	0.00
37 C	Benzo(a)pyrene	1.104	1.177	-6.6	117	0.00
38	Dibenzo(a,h)anthracene	1.112	1.129	-1.5	114	0.00
39	Benzo(a,h,i)perylene	1.149	1.195	-4.0	114	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061319\
Data File : BE100010.D
Acq On : 14 Jun 2019 3:24
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 14 04:47:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			