


n-Nitrosodimethylamine

Response Ratio




Resp Ratio = $4.03e-001 * \text{Amt} - 2.61e-001$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061719.M
Calibration Table Last Updated: Mon Jun 17 17:23:36 2019

bis(2-Chloroethyl)ether

Response Ratio




Resp Ratio = 1.26e+000 * Amt - 2.62e-001
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061719.M
Calibration Table Last Updated: Mon Jun 17 17:23:36 2019

Pentachlorophenol

Response Ratio



$R = 3.86e-003 A^2 + 6.50e-002 A - 1.29e-002$
Coef of Det (r^2) = 0.999 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061719.M
Calibration Table Last Updated: Mon Jun 17 17:23:36 2019