

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061719\
 Data File : BE100062.D
 Acq On : 17 Jun 2019 12:32
 Operator : JU/SJ
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC0.4

Manual Integrations
 APPROVED

mohammad
 6/20/2019 9:03:22 AM

Quant Time: Jun 17 13:11:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 17 13:03:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	695	0.40	ng	0.00
7) Naphthalene-d8	10.59	136	2342	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	1781	0.40	ng	0.00
19) Phenanthrene-d10	17.20	188	5658	0.40	ng	0.00
27) Chrysene-d12	21.39	240	8281	0.40	ng	0.00
34) Perylene-d12	23.92	264	9913	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	701	0.37	ng	0.00
5) Phenol-d6	6.96	99	858	0.36	ng	0.00
8) Nitrobenzene-d5	8.96	82	843	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.21	152	1640	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	378	0.29	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	2803	0.42	ng	0.00
25) Fluoranthene-d10	19.24	212	30496	0.37	ng	0.00
29) Terphenyl-d14	19.84	244	6041	0.41	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.26	88	481	0.468	ng	100
3) n-Nitrosodimethylamine	3.61	42	147	0.248	ng	# 100
6) bis(2-Chloroethyl)ether	7.22	93	714	0.362	ng	100
9) Naphthalene	10.64	128	10441	0.413	ng	100
10) Hexachlorobutadiene	10.90	225	896	0.465	ng	99
12) 2-Methylnaphthalene	12.28	142	1571	0.382	ng	100
16) Acenaphthylene	14.18	152	3503	0.395	ng	100
17) Acenaphthene	14.53	154	1940	0.402	ng	100
18) Fluorene	15.50	166	2849	0.393	ng	100
20) 4-Bromophenyl-phenylether	16.40	248	1320	0.412	ng	100
21) Hexachlorobenzene	16.51	284	1366	0.417	ng	100
22) Pentachlorophenol	16.89	266	362m	0.353	ng	
23) Phenanthrene	17.26	178	5309	0.387	ng	100
24) Anthracene	17.35	178	4520	0.359	ng	100
26) Fluoranthene	19.27	202	8765	0.376	ng	100
28) Pyrene	19.63	202	9083	0.417	ng	100
30) Benzo(a)anthracene	21.38	228	10188	0.412	ng	100
31) Chrysene	21.43	228	12047	0.471	ng	100
32) Bis(2-ethylhexyl)phthalate	21.29	149	13372	0.376	ng	100
33) Indeno(1,2,3-cd)pyrene	26.59	276	14131	0.466	ng	100
35) Benzo(b)fluoranthene	23.14	252	11609	0.418	ng	100
36) Benzo(k)fluoranthene	23.19	252	12713	0.438	ng	100
37) Benzo(a)pyrene	23.80	252	11429	0.418	ng	100
38) Dibenzo(a,h)anthracene	26.62	278	11129	0.404	ng	100
39) Benzo(g,h,i)perylene	27.42	276	11927	0.419	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061719\
Data File : BE100062.D
Acq On : 17 Jun 2019 12:32
Operator : JU/SJ
Sample : SSTDICCC0.4
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM

Quant Time: Jun 17 13:11:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 17 13:03:28 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.78 min Scan# 427
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
150	136.9	109.5	164.3
115	65.7	52.6	78.8

Manual Integrations
APPROVED




mohammad
6/20/2019 9:03:22 AM




#2

1,4-Dioxane
Concen: 0.468 ng
RT: 3.26 min Scan# 34
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
88	100		
58	52.4	41.9	62.9
43	19.8	15.8	23.8





#3





n-Nitrosodimethylamine
Concen: 0.248 ng
RT: 3.61 min Scan# 65
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	42	Resp:	147
Ion	Ratio	Lower	Upper
42	100		
74	0.0	0.0	0.0
44	0.0	0.0	0.0

Manual Integrations
APPROVED




mohammad
6/20/2019 9:03:22 AM






#4

2-Fluorophenol
Concen: 0.374 ng
RT: 5.35 min Scan# 216
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	112	Resp:	701
Ion	Ratio	Lower	Upper
112	100		
64	49.5	39.6	59.4
63	141.2	113.0	169.4





#5

Phenol-d6

Concen: 0.355 ng

RT: 6.96 min Scan# 356

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tot Ion:	99	Resp:	858
Ion	Ratio	Lower	Upper
99	100		
42	14.3	11.4	17.2
71	41.0	32.8	49.2

Instrument :





BNA_E

ClientSampleId :

SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



#6

bis(2-Chloroethyl)ether

Concen: 0.362 ng


RT: 7.22 min Scan# 378


Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tgt Ion:	93	Resp:	714
Ion	Ratio	Lower	Upper
93	100		
63	206.3	165.0	247.6
95	33.8	26.5	39.7





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.59 min Scan# 648

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E


ClientSampled :

SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
136	100		
137	15.8	12.6	19.0
54	41.3	33.0	49.6
68	10.7	8.6	12.8

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE1

Ion 68.00 (67.70 to 68.70): BE1



#8

Nitrobenzene-d5

Concen: 0.380 ng


RT: 8.96 min Scan# 523

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32



Tgt Ion	Ratio	Lower	Upper
82	100		
128	168.2	134.6	201.8
54	177.0	141.6	212.4




Abundance Ion 82.00 (81.70 to 82.70): BE1

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE1





#9

Naphthalene

Concen: 0.413 ng

RT: 10.64 min Scan# 652

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	128	Resp:	10441
Ion Ratio	Lower	Upper	
128	100		
129	3.7	3.0	4.4
127	4.2	3.4	5.0

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



#10

Hexachlorobutadiene

Concen: 0.465 ng




RT: 10.90 min Scan# 672


Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tgt Ion:	225	Resp:	896
Ion Ratio	Lower	Upper	
225	100		
223	62.9	51.1	76.7
227	64.3	51.9	77.9









#11
 2-Methylnaphthalene-d10
 Concen: 0.402 ng
 RT: 12.21 min Scan# 763
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
152	100		
151	19.3	15.4	23.2




Manual Integrations
 APPROVED


mohammad
 6/20/2019 9:03:22 AM



#12
 2-Methylnaphthalene
 Concen: 0.382 ng
 RT: 12.28 min Scan# 768
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.8	74.2	111.4
115	45.6	36.5	54.7





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.46 min Scan# 919

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:	164	Resp:	1781
Ion	Ratio	Lower	Upper
164	100		
162	97.0	77.6	116.4
160	52.1	41.7	62.5

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



#14

2,4,6-Tribromophenol

Concen: 0.287 ng




RT: 15.97 min Scan# 1026


Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tgt Ion:	330	Resp:	378
Ion	Ratio	Lower	Upper
330	100		
332	93.4	74.7	112.1
141	25.4	20.3	30.5









#15
2-Fluorobiphenyl
Concen: 0.415 ng
RT: 13.09 min Scan# 824
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampled :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
172	100		
171	38.8	31.0	46.6
170	26.6	21.3	31.9




Manual Integrations
APPROVED


mohammad
6/20/2019 9:03:22 AM



#16
Acenaphthylene
Concen: 0.395 ng
RT: 14.18 min Scan# 900
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.1	24.1
153	13.9	11.1	16.7






#17
Acenaphthene
Concen: 0.402 ng
RT: 14.53 min Scan# 924
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
154	100		
153	117.9	94.3	141.5
152	61.3	49.0	73.6

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18
Fluorene
Concen: 0.393 ng
RT: 15.50 min Scan# 991
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32




Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.9	82.3	123.5
167	13.7	11.0	16.4


Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.20 min Scan# 1118
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.3	8.2	12.4




Manual Integrations
APPROVED


mohammad
6/20/2019 9:03:22 AM



#20
4-Bromophenyl-phenylether
Concen: 0.412 ng
RT: 16.40 min Scan# 1058
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
248	100		
250	87.3	69.8	104.8
141	53.8	43.0	64.6





#21

Hexachlorobenzene

Concen: 0.417 ng

RT: 16.51 min Scan# 1066

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E


ClientSampleId :

SSTDICCC0.4




Tot Ion:	284	Resp:	1366
Ion Ratio		Lower	Upper
284	100		
142	25.5	20.3	30.5
249	31.6	25.3	37.9

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 0.353 ng m


RT: 16.89 min Scan# 1095

Delta R.T. 0.00 min



Lab File: BE100062.D


Acq: 17 Jun 2019 12:32

Tgt Ion:	266	Resp:	362
Ion Ratio		Lower	Upper
266	100		
264	85.7	68.6	102.8
268	83.8	67.0	100.6



Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 0.387 ng

RT: 17.26 min Scan# 1122

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:178	Resp:	5309
Ion Ratio	Lower	Upper
178	100	
176	19.6	15.7 23.5
179	15.5	12.4 18.6

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



#24

Anthracene

Concen: 0.359 ng




RT: 17.35 min Scan# 1129


Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tgt Ion:178	Resp:	4520
Ion Ratio	Lower	Upper
178	100	
176	18.4	14.6 21.8
179	15.7	12.6 19.0









#25
 Fluoranthene-d10
 Concen: 0.370 ng
 RT: 19.24 min Scan# 1388
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
212	100		
106	1.8	1.4	2.2
104	1.1	0.9	1.3




Manual Integrations
 APPROVED


mohammad
 6/20/2019 9:03:22 AM



#26
 Fluoranthene
 Concen: 0.376 ng
 RT: 19.27 min Scan# 1393
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
202	100		
101	6.6	5.3	7.9
203	16.6	13.3	19.9









#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.39 min Scan# 1631
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
240	100		
120	6.7	5.4	8.0
236	28.2	22.6	33.8




Manual Integrations
APPROVED


mohammad
6/20/2019 9:03:22 AM



#28
Pyrene
Concen: 0.417 ng
RT: 19.63 min Scan# 1456
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
202	100		
200	19.5	15.6	23.4
203	17.5	14.0	21.0









#29
 Terphenyl-d14
 Concen: 0.405 ng
 RT: 19.84 min Scan# 1492
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
244	100		
212	34.9	27.9	41.9
122	7.2	5.8	8.6




Manual Integrations
 APPROVED


mohammad
 6/20/2019 9:03:22 AM



#30
 Benzo(a)anthracene
 Concen: 0.412 ng
 RT: 21.38 min Scan# 1630
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.1	33.1
229	21.1	16.9	25.3





#31

Chrysene

Concen: 0.471 ng

RT: 21.43 min Scan# 1634

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Instrument :

BNA_E





ClientSampleId :

SSTDICCC0.4

Tot Ion:228	Resp:	12047
Ion Ratio	Lower	Upper
228 100		
226 30.1	24.1	36.1
229 19.8	15.8	23.8

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.376 ng




RT: 21.29 min Scan# 1623


Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tgt Ion:149	Resp:	13372
Ion Ratio	Lower	Upper
149 100		
167 7.4	5.9	8.9
279 1.3	1.0	1.6









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.466 ng
 RT: 26.59 min Scan# 2958
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.4

Tot Ion	Ratio	Lower	Upper
276	100		
138	13.4	10.7	16.1
277	23.9	19.0	28.6




Manual Integrations
 APPROVED


mohammad
 6/20/2019 9:03:22 AM



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.92 min Scan# 2213
 Delta R.T. 0.00 min
 Lab File: BE100062.D
 Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.8	20.6	31.0
265	29.5	23.6	35.4









#35
Benzo(b)fluoranthene
Concen: 0.418 ng
RT: 23.14 min Scan# 1993
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.5	19.6	29.4
125	9.0	7.2	10.8




Manual Integrations
APPROVED


mohammad
6/20/2019 9:03:22 AM



#36
Benzo(k)fluoranthene
Concen: 0.438 ng
RT: 23.19 min Scan# 2008
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	19.3	28.9
125	7.0	5.6	8.4









#37
Benzo(a)pyrene
Concen: 0.418 ng
RT: 23.80 min Scan# 2180
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Instrument :
BNA_E
ClientSampleId :
SSTDICCC0.4

Tot Ion	Ratio	Lower	Upper
252	100		
253	25.0	20.0	30.0
125	12.0	9.6	14.4




Manual Integrations
APPROVED




mohammad
6/20/2019 9:03:22 AM



#38
Dibenzo(a,h)anthracene
Concen: 0.404 ng
RT: 26.62 min Scan# 2965
Delta R.T. 0.00 min
Lab File: BE100062.D
Acq: 17 Jun 2019 12:32

Tgt Ion	Ratio	Lower	Upper
278	100		
139	11.5	9.2	13.8
279	26.7	21.4	32.0





#39

Benzo(a,h,i)perylene

Concen: 0.419 ng

RT: 27.42 min Scan# 3189

Delta R.T. 0.00 min

Lab File: BE100062.D

Acq: 17 Jun 2019 12:32

Tot Ion:	276	Resp:	11927
Ion Ratio		Lower	Upper
276	100		
277	24.3	19.4	29.2
138	13.8	11.0	16.6

Instrument :

BNA_E

ClientSampleId :

SSTDICCC0.4

Manual Integrations
APPROVED

mohammad
6/20/2019 9:03:22 AM

