

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
 Data File : BE096806.D
 Acq On : 18 Jun 2018 13:17
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE061818

Quant Time: Jun 18 14:42:29 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 18 13:28:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.648	0.645	0.5	94	0.00
3	n-Nitrosodimethylamine	0.683	0.650	4.8	95	0.00
4 S	2-Fluorophenol	0.874	0.855	2.2	101	0.00
5 S	Phenol-d6	1.310	1.284	2.0	104	0.00
6	bis(2-Chloroethyl)ether	1.545	1.427	7.6	93	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
8 S	Nitrobenzene-d5	0.222	0.217	2.3	101	0.00
9	Naphthalene	3.930	3.715	5.5	93	0.00
10	Hexachlorobutadiene	0.164	0.157	4.3	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.599	6.3	93	0.00
12	2-Methylnaphthalene	0.659	0.613	7.0	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.083	0.080	3.6	109	0.01
15 S	2-Fluorobiphenyl	1.645	1.565	4.9	94	0.00
16	Acenaphthylene	1.887	1.690	10.4	94	0.00
17	Acenaphthene	1.268	1.181	6.9	94	0.00
18	Fluorene	1.500	1.383	7.8	94	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.195	0.180	7.7	94	0.00
21	Hexachlorobenzene	0.229	0.211	7.9	93	0.00
22	Pentachlorophenol	0.036	0.034	5.6	109	0.00
23	Phenanthrene	1.087	1.018	6.3	94	0.00
24	Anthracene	0.892	0.814	8.7	98	0.00
25 SURR	Fluoranthene-d10	3.915	3.547	9.4	96	0.00
26	Fluoranthene	1.151	1.031	10.4	94	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
28	Pvrene	1.314	1.242	5.5	96	0.00
29 S	Terphenyl-d14	0.906	0.864	4.6	95	0.00
30	Benzo(a)anthracene	0.970	0.850	12.4	96	0.00
31	Chrysene	1.221	1.147	6.1	95	0.00
32	Bis(2-ethylhexyl)phthalate	0.729	0.692	5.1	106	0.00
33	Indeno(1,2,3-cd)pyrene	0.917	0.866	5.6	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.099	0.932	15.2	98	0.00
36	Benzo(k)fluoranthene	1.307	1.149	12.1	101	0.00
37 C	Benzo(a)pyrene	0.954	0.777	18.6	96	0.00
38	Dibenzo(a,h)anthracene	0.916	0.801	12.6	104	0.00
39	Benzo(a,h,i)perylene	1.014	0.888	12.4	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
Data File : BE096806.D
Acq On : 18 Jun 2018 13:17
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE061818

Quant Time: Jun 18 14:42:29 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 18 13:28:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			