

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
 Data File : BE096808.D
 Acq On : 18 Jun 2018 14:52
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 18 16:02:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 18 13:28:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.648	0.659	-1.7	99	0.00
3	n-Nitrosodimethylamine	0.683	0.628	8.1	95	0.00
4 S	2-Fluorophenol	0.874	0.876	-0.2	107	0.00
5 S	Phenol-d6	1.310	1.294	1.2	108	0.00
6	bis(2-Chloroethyl)ether	1.545	1.438	6.9	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.222	0.211	5.0	101	0.00
9	Naphthalene	3.930	3.733	5.0	96	0.00
10	Hexachlorobutadiene	0.164	0.152	7.3	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.608	4.9	96	0.00
12	2-Methylnaphthalene	0.659	0.621	5.8	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.083	0.081	2.4	117	0.00
15 S	2-Fluorobiphenyl	1.645	1.509	8.3	96	0.00
16	Acenaphthylene	1.887	1.648	12.7	97	0.00
17	Acenaphthene	1.268	1.147	9.5	96	0.00
18	Fluorene	1.500	1.358	9.5	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.195	0.175	10.3	97	0.00
21	Hexachlorobenzene	0.229	0.208	9.2	97	0.00
22	Pentachlorophenol	0.036	0.033	8.3	113	0.00
23	Phenanthrene	1.087	0.979	9.9	96	0.00
24	Anthracene	0.892	0.786	11.9	101	0.00
25 SURR	Fluoranthene-d10	3.915	3.468	11.4	100	0.00
26	Fluoranthene	1.151	1.026	10.9	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.314	1.204	8.4	100	0.00
29 S	Terphenyl-d14	0.906	0.860	5.1	102	0.00
30	Benzo(a)anthracene	0.970	0.858	11.5	104	0.00
31	Chrysene	1.221	1.142	6.5	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.729	0.684	6.2	112	0.00
33	Indeno(1,2,3-cd)pyrene	0.917	0.772	15.8	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.099	1.006	8.5	106	0.00
36	Benzo(k)fluoranthene	1.307	1.163	11.0	102	0.00
37 C	Benzo(a)pyrene	0.954	0.810	15.1	100	0.00
38	Dibenzo(a,h)anthracene	0.916	0.794	13.3	103	0.01
39	Benzo(a,h,i)perylene	1.014	0.913	10.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
Data File : BE096808.D
Acq On : 18 Jun 2018 14:52
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 18 16:02:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 18 13:28:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			