

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
 Data File : BE096808.D
 Acq On : 18 Jun 2018 14:52
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 18 16:02:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 18 13:28:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	99	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	95	0.00
4 S	2-Fluorophenol	0.400	0.401	-0.3	107	0.00
5 S	Phenol-d6	0.400	0.395	1.3	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.372	7.0	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.379	5.3	101	0.00
9	Naphthalene	0.400	0.380	5.0	96	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.380	5.0	96	0.00
12	2-Methylnaphthalene	0.400	0.377	5.8	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.387	3.3	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.367	8.3	96	0.00
16	Acenaphthylene	0.400	0.349	12.8	97	0.00
17	Acenaphthene	0.400	0.362	9.5	96	0.00
18	Fluorene	0.400	0.362	9.5	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.359	10.3	97	0.00
21	Hexachlorobenzene	0.400	0.363	9.3	97	0.00
22	Pentachlorophenol	0.400	0.373	6.8	113	0.00
23	Phenanthrene	0.400	0.360	10.0	96	0.00
24	Anthracene	0.400	0.352	12.0	101	0.00
25 SURR	Fluoranthene-d10	0.400	0.354	11.5	100	0.00
26	Fluoranthene	0.400	0.357	10.8	100	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
28	Pvrene	0.400	0.367	8.3	100	0.00
29 S	Terphenyl-d14	0.400	0.380	5.0	102	0.00
30	Benzo(a)anthracene	0.400	0.354	11.5	104	0.00
31	Chrysene	0.400	0.374	6.5	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.376	6.0	112	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.337	15.8	99	0.00
34 I	Perylene-d12	0.400	0.400	0.0	102	0.00
35	Benzo(b)fluoranthene	0.400	0.366	8.5	106	0.00
36	Benzo(k)fluoranthene	0.400	0.356	11.0	102	0.00
37 C	Benzo(a)pyrene	0.400	0.340	15.0	100	0.00
38	Dibenzo(a,h)anthracene	0.400	0.357	10.8	103	0.01
39	Benzo(a,h,i)perylene	0.400	0.360	10.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
Data File : BE096808.D
Acq On : 18 Jun 2018 14:52
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 18 16:02:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 18 13:28:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			