

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
 Data File : BE096825.D
 Acq On : 19 Jun 2018 1:25
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 19 02:17:59 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 18 13:28:13 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.648	0.575	11.3	95	-0.01
3	n-Nitrosodimethylamine	0.683	0.697	-2.0	117	0.00
4 S	2-Fluorophenol	0.874	1.165	-33.3#	157#	-0.01
5 S	Phenol-d6	1.310	1.791	-36.7#	165#	0.00
6	bis(2-Chloroethyl)ether	1.545	1.535	0.6	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.222	0.283	-27.5#	156#	0.00
9	Naphthalene	3.930	3.666	6.7	109	0.00
10	Hexachlorobutadiene	0.164	0.150	8.5	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.602	5.8	110	0.00
12	2-Methylnaphthalene	0.659	0.612	7.1	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.083	0.132	-59.0#	220#	0.01
15 S	2-Fluorobiphenyl	1.645	1.435	12.8	105	0.00
16	Acenaphthylene	1.887	1.779	5.7	120	0.00
17	Acenaphthene	1.268	1.156	8.8	112	0.00
18	Fluorene	1.500	1.393	7.1	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
20	4-Bromophenyl-phenylether	0.195	0.180	7.7	109	0.00
21	Hexachlorobenzene	0.229	0.194	15.3	99	0.00
22	Pentachlorophenol	0.036	0.067	-86.1#	252#	0.00
23	Phenanthrene	1.087	0.966	11.1	104	0.00
24	Anthracene	0.892	0.850	4.7	119	0.00
25 SURR	Fluoranthene-d10	3.915	3.862	1.4	122	0.00
26	Fluoranthene	1.151	1.094	5.0	117	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
28	Pvrene	1.314	1.166	11.3	125	0.00
29 S	Terphenyl-d14	0.906	0.793	12.5	121	0.00
30	Benzo(a)anthracene	0.970	1.051	-8.4	165#	0.00
31	Chrysene	1.221	1.087	11.0	125	0.00
32	Bis(2-ethylhexyl)phthalate	0.729	1.600	-119.5#	339#	0.01
33	Indeno(1,2,3-cd)pyrene	0.917	1.096	-19.5	182#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	159#	0.00
35	Benzo(b)fluoranthene	1.099	1.014	7.7	166#	0.00
36	Benzo(k)fluoranthene	1.307	1.059	19.0	145	0.00
37 C	Benzo(a)pyrene	0.954	0.949	0.5	184#	0.00
38	Dibenzo(a,h)anthracene	0.916	0.895	2.3	181#	0.00
39	Benzo(a,h,i)perylene	1.014	0.933	8.0	164#	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE061818\
Data File : BE096825.D
Acq On : 19 Jun 2018 1:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 19 02:17:59 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 18 13:28:13 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			