

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE061919\
 Data File : BE100088.D
 Acq On : 19 Jun 2019 18:38
 Operator : JU/SJ
 Sample : K3309-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 LF001MW001-190610

Quant Time: Jun 20 03:50:29 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	722	0.40	ng	0.00
7) Naphthalene-d8	10.54	136	2518	0.40	ng	0.00
13) Acenaphthene-d10	14.41	164	5362	0.40	ng	0.00
19) Phenanthrene-d10	17.16	188	7512	0.40	ng	0.00
27) Chrysene-d12	21.34	240	10041	0.40	ng	0.00
34) Perylene-d12	23.84	264	10584	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.30	112	578	0.39	ng	0.00
5) Phenol-d6	6.90	99	1513	0.74	ng	-0.03
8) Nitrobenzene-d5	8.89	82	1126	0.46	ng	-0.03
11) 2-Methylnaphthalene-d10	12.14	152	1740	0.36	ng	-0.02
14) 2,4,6-Tribromophenol	15.90	330	754	0.25	ng	-0.01
15) 2-Fluorobiphenyl	13.03	172	3598	0.17	ng	-0.02
25) Fluoranthene-d10	19.20	212	32597	0.30	ng	0.00
29) Terphenyl-d14	19.79	244	9476	0.47	ng	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.19	88	287	0.248	ng	# 1
6) bis(2-Chloroethyl)ether	7.09	93	940	0.476	ng	# 1
9) Naphthalene	10.59	128	2634961	96.298	ng	# 97
12) 2-Methylnaphthalene	12.21	142	69148	15.367	ng	96
16) Acenaphthylene	14.12	152	642	0.025	ng	# 22
17) Acenaphthene	14.47	154	416	0.029	ng	# 55
18) Fluorene	15.45	166	1721	0.081	ng	94
22) Pentachlorophenol	16.81	266	54	0.276	ng	96
23) Phenanthrene	17.20	178	1267	0.070	ng	# 74
28) Pyrene	19.56	202	870	0.034	ng	# 54
30) Benzo(a)anthracene	21.33	228	1424	0.048	ng	# 1
31) Chrysene	21.33	228	1008	0.029	ng	# 1
32) Bis(2-ethylhexyl)phthalate	21.24	149	4126	0.146	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE061919\
Data File : BE100088.D
Acq On : 19 Jun 2019 18:38
Operator : JU/SJ
Sample : K3309-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
LF001MW001-190610

Quant Time: Jun 20 03:50:29 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.73 min Scan# 422
Delta R.T. -0.01 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38

Instrument :
BNA_E
ClientSampleId :
LF001MW001-190610




Tgt Ion: 152 Resp: 722
Ion Ratio Lower Upper
152 100
150 159.7 99.8 149.8#
115 101.5 51.3 76.9#




#2

1,4-Dioxane
Concen: 0.248 ng
RT: 3.19 min Scan# 28
Delta R.T. -0.01 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38

Tgt Ion: 88 Resp: 287
Ion Ratio Lower Upper
88 100
58 104.2 51.1 76.7#
43 391.6 21.7 32.5#





#4

2-Fluorophenol

Concen: 0.386 ng

RT: 5.30 min Scan# 211

Delta R.T. -0.01 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion: 112 Resp: 578

Ion Ratio Lower Upper

112 100

64 0.0 22.0 33.0#




63 0.0 121.8 182.6#



Abundance Ion 112.00 (111.70 to 112.70): BE1

Ion 64.00 (63.70 to 64.70): BE1

Ion 63.00 (62.70 to 63.70): BE1



#5

Phenol-d6

Concen: 0.735 ng

RT: 6.90 min Scan# 350

Delta R.T. -0.03 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion: 99 Resp: 1513

Ion Ratio Lower Upper

99 100

42 0.0 0.0 0.0



71 0.0 37.6 56.4#




Abundance Ion 99.00 (98.70 to 99.70): BE1

Ion 42.00 (41.70 to 42.70): BE1

Ion 71.00 (70.70 to 71.70): BE1





#6

bis(2-Chloroethyl)ether

Concen: 0.476 ng

RT: 7.09 min Scan# 367

Delta R.T. -0.10 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610





Tgt Ion: 93 Resp: 940

Ion Ratio Lower Upper

93 100

63 3095.9 200.0 300.0#

95 0.0 24.0 36.0#



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.54 min Scan# 644

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Tgt Ion: 136 Resp: 2518




Ion Ratio Lower Upper


136 100

137 16.2 11.6 17.4

54 64.8 33.9 50.9#

68 17.4 7.8 11.8#





#8

Nitrobenzene-d5

Concen: 0.461 ng

RT: 8.89 min Scan# 517

Delta R.T. -0.03 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion: 82 Resp: 1126

Ion Ratio Lower Upper

82 100

128 264.6 128.0 192.0#




54 179.1 143.4 215.0



Abundance Ion 82.00 (81.70 to 82.70): BE1

Ion 128.00 (127.70 to 128.70): BE1

Ion 54.00 (53.70 to 54.70): BE1



#9

Naphthalene

Concen: 96.298 ng

RT: 10.59 min Scan# 648

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion: 128 Resp: 2634961

Ion Ratio Lower Upper

128 100

129 2.8 3.1 4.7#



127 3.3 3.4 5.2#




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#11
 2-Methylnaphthalene-d10
 Concen: 0.363 ng
 RT: 12.14 min Scan# 758
 Delta R.T. -0.02 min
 Lab File: BE100088.D
 Acq: 19 Jun 2019 18:38




Instrument :
 BNA_E
 ClientSampleId :
 LF001MW001-190610


Tgt Ion:152 Resp: 1740
 Ion Ratio Lower Upper
 152 100
 151 0.0 16.3 24.5#



#12
 2-Methylnaphthalene
 Concen: 15.367 ng
 RT: 12.21 min Scan# 763
 Delta R.T. -0.02 min
 Lab File: BE100088.D
 Acq: 19 Jun 2019 18:38

Tgt Ion:142 Resp: 69148
 Ion Ratio Lower Upper
 142 100
 141 90.7 74.1 111.1
 115 45.4 40.9 61.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.41 min Scan# 916
Delta R.T. -0.00 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38




Instrument :
BNA_E
ClientSampleId :
LF001MW001-190610


Tgt Ion:164 Resp: 5362
Ion Ratio Lower Upper
164 100
162 49.7 78.2 117.2#
160 106.3 45.6 68.4#



#14
2,4,6-Tribromophenol
Concen: 0.255 ng
RT: 15.90 min Scan# 1021
Delta R.T. -0.01 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38

Tgt Ion:330 Resp: 754
Ion Ratio Lower Upper
330 100
332 99.3 73.3 109.9
141 27.9 22.2 33.2









#15
2-Fluorobiphenyl
Concen: 0.171 ng
RT: 13.03 min Scan# 820
Delta R.T. -0.02 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38




Instrument :
BNA_E
ClientSampleId :
LF001MW001-190610


Tgt Ion:172 Resp: 3598
Ion Ratio Lower Upper
172 100
171 38.6 29.8 44.8
170 27.6 23.2 34.8



#16
Acenaphthylene
Concen: 0.025 ng
RT: 14.12 min Scan# 896
Delta R.T. -0.02 min
Lab File: BE100088.D
Acq: 19 Jun 2019 18:38

Tgt Ion:152 Resp: 642
Ion Ratio Lower Upper
152 100
151 26.6 15.9 23.9#
153 83.0 10.1 15.1#





#17

Acenaphthene

Concen: 0.029 ng

RT: 14.47 min Scan# 920

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion:154 Resp: 416

Ion Ratio Lower Upper

154 100

153 94.2 95.5 143.3#




152 0.0 54.9 82.3#



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.081 ng

RT: 15.45 min Scan# 987

Delta R.T. -0.01 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion:166 Resp: 1721

Ion Ratio Lower Upper

166 100

165 106.8 79.8 119.8



167 12.7 10.2 15.4




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.16 min Scan# 1115

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion:188 Resp: 7512

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0




80 11.6 6.3 9.5#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE1

Ion 80.00 (79.70 to 80.70): BE1



#22

Pentachlorophenol

Concen: 0.276 ng

RT: 16.81 min Scan# 1089

Delta R.T. -0.05 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion:266 Resp: 54

Ion Ratio Lower Upper

266 100

264 80.9 67.8 101.6



268 91.0 70.0 105.0




Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 0.070 ng

RT: 17.20 min Scan# 1118

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion:178 Resp: 1267

Ion Ratio Lower Upper

178 100

176 0.0 15.9 23.9#




179 14.5 13.1 19.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#25

Fluoranthene-d10

Concen: 0.304 ng

RT: 19.20 min Scan# 1380

Delta R.T. 0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion:212 Resp: 32597

Ion Ratio Lower Upper

212 100

106 2.0 1.4 2.0



104 1.4 0.8 1.2#




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E





#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.34 min Scan# 1628

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion:240 Resp: 10041

Ion Ratio Lower Upper

240 100

120 10.3 6.6 10.0#




236 31.1 23.8 35.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#28

Pyrene

Concen: 0.034 ng

RT: 19.56 min Scan# 1444

Delta R.T. -0.03 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion:202 Resp: 870

Ion Ratio Lower Upper

202 100

200 29.8 15.6 23.4#



203 49.0 13.9 20.9#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#29

Terphenyl-d14

Concen: 0.472 ng

RT: 19.79 min Scan# 1484

Delta R.T. -0.01 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610


Tgt Ion:244 Resp: 9476

Ion Ratio Lower Upper

244 100

212 40.1 29.8 44.8




122 10.4 6.1 9.1#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#30

Benzo(a)anthracene

Concen: 0.048 ng

RT: 21.33 min Scan# 1627

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38


Tgt Ion:228 Resp: 1424

Ion Ratio Lower Upper

228 100

226 178.6 23.0 34.4#



229 62.3 16.4 24.6#




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#31

Chrysene

Concen: 0.029 ng

RT: 21.33 min Scan# 1627

Delta R.T. -0.06 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610





Tgt Ion: 228 Resp: 1008

Ion Ratio Lower Upper

228 100

226 178.6 23.8 35.8#

229 62.3 16.3 24.5#



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.146 ng

RT: 21.24 min Scan# 1620

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38




Tgt Ion: 149 Resp: 4126




Ion Ratio Lower Upper

149 100

167 8.5 5.8 8.6

279 1.6 1.1 1.7





#34

Perylene-d12

Concen: 0.400 ng

RT: 23.84 min Scan# 2192

Delta R.T. -0.00 min

Lab File: BE100088.D

Acq: 19 Jun 2019 18:38

Instrument :

BNA_E

ClientSampleId :

LF001MW001-190610

Tgt Ion:264 Resp: 10584

Ion Ratio Lower Upper

264 100

260 26.0 20.8 31.2

265 29.4 23.2 34.8

