


n-Nitrosodimethylamine

Response Ratio




Resp Ratio = $4.21 \times 10^{-1} \times \text{Amt} - 3.59 \times 10^{-1}$
Coef of Det (r^2) = 0.994 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Calibration Table Last Updated: Wed Jun 19 15:07:24 2019


bis(2-Chloroethyl)ether

Response Ratio



Resp Ratio = 1.18×10^0 * Amt - 9.88×10^{-2}
Coef of Det (r^2) = 0.997 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Calibration Table Last Updated: Wed Jun 19 15:07:24 2019



Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Calibration Table Last Updated: Wed Jun 19 15:07:24 2019