

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061919\
 Data File : BE100099.D
 Acq On : 20 Jun 2019 1:51
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 20 03:21:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	133	-0.01
2	1,4-Dioxane	0.400	0.374	6.5	141	0.00
3	n-Nitrosodimethylamine	0.400	0.517	-29.3#	351	-0.03
4 S	2-Fluorophenol	0.400	0.557	-39.3#	197	0.00
5 S	Phenol-d6	0.400	0.480	-20.0	158	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	143	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.421	-5.2	139	-0.01
9	Naphthalene	0.400	0.405	-1.3	135	-0.01
10	Hexachlorobutadiene	0.400	0.370	7.5	126	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	139	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	145	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	141	0.00
14 S	2,4,6-Tribromophenol	0.400	0.518	-29.5#	205	0.00
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	136	-0.01
16	Acenaphthylene	0.400	0.399	0.3	139	0.00
17	Acenaphthene	0.400	0.406	-1.5	140	0.00
18	Fluorene	0.400	0.407	-1.7	141	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	144	0.00
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	148	0.00
21	Hexachlorobenzene	0.400	0.382	4.5	139	0.00
22	Pentachlorophenol	0.400	0.599	-49.7#	285	-0.04
23	Phenanthrene	0.400	0.385	3.8	141	0.00
24	Anthracene	0.400	0.414	-3.5	161	0.00
25 SURR	Fluoranthene-d10	0.400	0.396	1.0	145	0.00
26	Fluoranthene	0.400	0.395	1.3	144	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	136	0.00
28	Pvrene	0.400	0.435	-8.7	150	0.00
29 S	Terphenyl-d14	0.400	0.441	-10.2	150	0.00
30	Benzo(a)anthracene	0.400	0.414	-3.5	154	0.00
31	Chrysene	0.400	0.396	1.0	131	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.581	-45.2#	229	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.400	0.0	131	0.00
34 I	Perylene-d12	0.400	0.400	0.0	138	0.00
35	Benzo(b)fluoranthene	0.400	0.369	7.8	128	0.00
36	Benzo(k)fluoranthene	0.400	0.363	9.3	129	0.00
37 C	Benzo(a)pyrene	0.400	0.382	4.5	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.371	7.3	132	0.00
39	Benzo(a,h,i)perylene	0.400	0.376	6.0	131	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061919\
Data File : BE100099.D
Acq On : 20 Jun 2019 1:51
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 20 03:21:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				