

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061919\
 Data File : BE100099.D
 Acq On : 20 Jun 2019 1:51
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 20 03:21:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	-0.01
2	1,4-Dioxane	0.642	0.599	6.7	141	0.00
3	n-Nitrosodimethylamine	0.238	0.185	22.3	351#	-0.03
4 S	2-Fluorophenol	0.830	1.156	-39.3#	197#	0.00
5 S	Phenol-d6	1.140	1.366	-19.8	158#	-0.01
6	bis(2-Chloroethyl)ether	1.119	1.071	4.3	143	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
8 S	Nitrobenzene-d5	0.388	0.409	-5.4	139	-0.01
9	Naphthalene	4.347	4.396	-1.1	135	-0.01
10	Hexachlorobutadiene	0.427	0.395	7.5	126	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.761	0.0	139	0.00
12	2-Methylnaphthalene	0.715	0.744	-4.1	145	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	141	0.00
14 S	2,4,6-Tribromophenol	0.221	0.286	-29.4#	205#	0.00
15 S	2-Fluorobiphenyl	1.573	1.643	-4.5	136	-0.01
16	Acenaphthylene	1.899	1.893	0.3	139	0.00
17	Acenaphthene	1.062	1.078	-1.5	140	0.00
18	Fluorene	1.581	1.607	-1.6	141	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	144	0.00
20	4-Bromophenyl-phenylether	0.238	0.236	0.8	148	0.00
21	Hexachlorobenzene	0.262	0.251	4.2	139	0.00
22	Pentachlorophenol	0.053	0.076	-43.4#	285#	-0.04
23	Phenanthrene	0.958	0.922	3.8	141	0.00
24	Anthracene	0.808	0.836	-3.5	161#	0.00
25 SURR	Fluoranthene-d10	5.718	5.661	1.0	145	0.00
26	Fluoranthene	1.538	1.517	1.4	144	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
28	Pvrene	1.028	1.117	-8.7	150	0.00
29 S	Terphenyl-d14	0.800	0.881	-10.1	150	0.00
30	Benzo(a)anthracene	1.190	1.233	-3.6	154#	0.00
31	Chrysene	1.383	1.371	0.9	131	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.632	-45.3#	229#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.556	0.0	131	0.00
34 I	Perylene-d12	1.000	1.000	0.0	138	0.00
35	Benzo(b)fluoranthene	1.174	1.084	7.7	128	0.00
36	Benzo(k)fluoranthene	1.335	1.211	9.3	129	0.00
37 C	Benzo(a)pyrene	1.133	1.081	4.6	132	0.00
38	Dibenzo(a,h)anthracene	1.161	1.078	7.1	132	0.00
39	Benzo(a,h,i)perylene	1.207	1.135	6.0	131	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061919\
Data File : BE100099.D
Acq On : 20 Jun 2019 1:51
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 20 03:21:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			