

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
 Data File : BE100102.D
 Acq On : 20 Jun 2019 10:48
 Operator : JU/SJ
 Sample : PB120562BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB120562BL

Quant Time: Jun 21 14:14:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	475	0.40	ng	0.01
7) Naphthalene-d8	10.57	136	1667	0.40	ng	0.04
13) Acenaphthene-d10	14.43	164	1399	0.40	ng	0.01
19) Phenanthrene-d10	17.19	188	3876	0.40	ng	0.03
27) Chrysene-d12	21.35	240	5520	0.40	ng	0.01
34) Perylene-d12	23.84	264	7857	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.32	112	323	0.33	ng	0.01
5) Phenol-d6	7.06	99	316	0.23	ng	0.13
8) Nitrobenzene-d5	9.03	82	306	0.19	ng	0.12
11) 2-Methylnaphthalene-d10	12.21	152	912	0.29	ng	0.06
14) 2,4,6-Tribromophenol	15.98	330	98	0.13	ng	0.07
15) 2-Fluorobiphenyl	13.07	172	1718	0.31	ng	0.03
25) Fluoranthene-d10	19.21	212	14924	0.27	ng	0.02
29) Terphenyl-d14	19.80	244	3761	0.34	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
Data File : BE100102.D
Acq On : 20 Jun 2019 10:48
Operator : JU/SJ
Sample : PB120562BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB120562BL

Quant Time: Jun 21 14:14:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

