

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
 Data File : BE100101.D
 Acq On : 20 Jun 2019 10:11
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 21 01:53:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.01
2	1,4-Dioxane	0.642	0.620	3.4	97	-0.01
3	n-Nitrosodimethylamine	0.238	0.231	2.9	292#	-0.02
4 S	2-Fluorophenol	0.830	0.833	-0.4	94	-0.01
5 S	Phenol-d6	1.140	1.220	-7.0	94	0.00
6	bis(2-Chloroethyl)ether	1.119	1.239	-10.7	110	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.388	0.360	7.2	88	0.01
9	Naphthalene	4.347	4.400	-1.2	97	0.00
10	Hexachlorobutadiene	0.427	0.410	4.0	94	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.775	-1.8	101	0.00
12	2-Methylnaphthalene	0.715	0.721	-0.8	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.221	0.231	-4.5	132	0.00
15 S	2-Fluorobiphenyl	1.573	1.628	-3.5	107	0.00
16	Acenaphthylene	1.899	1.936	-1.9	113	-0.01
17	Acenaphthene	1.062	1.117	-5.2	115	0.00
18	Fluorene	1.581	1.635	-3.4	114	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.238	0.241	-1.3	127	0.00
21	Hexachlorobenzene	0.262	0.246	6.1	113	-0.01
22	Pentachlorophenol	0.053	0.055	-3.8	174#	-0.01
23	Phenanthrene	0.958	0.927	3.2	118	0.00
24	Anthracene	0.808	0.791	2.1	127	0.00
25 SURR	Fluoranthene-d10	5.718	5.952	-4.1	127	0.00
26	Fluoranthene	1.538	1.584	-3.0	126	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	127	0.00
28	Pvrene	1.028	1.050	-2.1	131	0.00
29 S	Terphenyl-d14	0.800	0.830	-3.7	132	-0.01
30	Benzo(a)anthracene	1.190	1.219	-2.4	142	0.00
31	Chrysene	1.383	1.390	-0.5	124	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.517	-35.1#	199#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.653	-6.2	130	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	137	-0.01
35	Benzo(b)fluoranthene	1.174	1.067	9.1	126	0.00
36	Benzo(k)fluoranthene	1.335	1.275	4.5	135	-0.01
37 C	Benzo(a)pyrene	1.133	1.117	1.4	135	-0.01
38	Dibenzo(a,h)anthracene	1.161	1.093	5.9	133	0.00
39	Benzo(a,h,i)perylene	1.207	1.123	7.0	129	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
Data File : BE100101.D
Acq On : 20 Jun 2019 10:11
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 21 01:53:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			