

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062019\
 Data File : BE100111.D
 Acq On : 20 Jun 2019 17:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 21 02:06:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	-0.01
2	1,4-Dioxane	0.642	0.728	-13.4	105	-0.01
3	n-Nitrosodimethylamine	0.238	0.184	22.7	214#	-0.02
4 S	2-Fluorophenol	0.830	1.014	-22.2	105	-0.01
5 S	Phenol-d6	1.140	1.205	-5.7	85	-0.01
6	bis(2-Chloroethyl)ether	1.119	1.198	-7.1	98	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.388	0.358	7.7	80	0.00
9	Naphthalene	4.347	4.239	2.5	86	0.00
10	Hexachlorobutadiene	0.427	0.419	1.9	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.766	-0.7	92	0.00
12	2-Methylnaphthalene	0.715	0.729	-2.0	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.221	0.243	-10.0	126	0.00
15 S	2-Fluorobiphenyl	1.573	1.550	1.5	93	0.00
16	Acenaphthylene	1.899	1.871	1.5	100	-0.01
17	Acenaphthene	1.062	1.064	-0.2	100	0.00
18	Fluorene	1.581	1.621	-2.5	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.238	0.240	-0.8	112	0.00
21	Hexachlorobenzene	0.262	0.250	4.6	103	-0.01
22	Pentachlorophenol	0.053	0.061	-15.1	170#	-0.03
23	Phenanthrene	0.958	0.941	1.8	107	0.00
24	Anthracene	0.808	0.792	2.0	113	0.00
25 SURR	Fluoranthene-d10	5.718	5.892	-3.0	112	0.00
26	Fluoranthene	1.538	1.581	-2.8	111	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pyrene	1.028	1.054	-2.5	115	0.00
29 S	Terphenyl-d14	0.800	0.829	-3.6	115	0.00
30	Benzo(a)anthracene	1.190	1.188	0.2	120	0.00
31	Chrysene	1.383	1.418	-2.5	110	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.435	-27.8#	164#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.601	-2.9	110	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.174	1.071	8.8	106	0.00
36	Benzo(k)fluoranthene	1.335	1.281	4.0	114	0.00
37 C	Benzo(a)pyrene	1.133	1.070	5.6	109	0.00
38	Dibenzo(a,h)anthracene	1.161	1.099	5.3	113	0.00
39	Benzo(g,h,i)perylene	1.207	1.158	4.1	112	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062019\
Data File : BE100111.D
Acq On : 20 Jun 2019 17:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 21 02:06:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		