

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062019\
 Data File : BE100111.D
 Acq On : 20 Jun 2019 17:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 21 02:06:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	81	-0.01
2	1,4-Dioxane	0.400	0.454	-13.5	105	-0.01
3	n-Nitrosodimethylamine	0.400	0.516	-29.0#	214	-0.02
4 S	2-Fluorophenol	0.400	0.489	-22.2	105	-0.01
5 S	Phenol-d6	0.400	0.423	-5.7	85	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.441	-10.2	98	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	88	0.00
8 S	Nitrobenzene-d5	0.400	0.369	7.8	80	0.00
9	Naphthalene	0.400	0.390	2.5	86	0.00
10	Hexachlorobutadiene	0.400	0.393	1.8	89	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.403	-0.8	92	0.00
12	2-Methylnaphthalene	0.400	0.408	-2.0	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.440	-10.0	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	93	0.00
16	Acenaphthylene	0.400	0.394	1.5	100	-0.01
17	Acenaphthene	0.400	0.401	-0.3	100	0.00
18	Fluorene	0.400	0.410	-2.5	103	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.400	0.404	-1.0	112	0.00
21	Hexachlorobenzene	0.400	0.381	4.8	103	-0.01
22	Pentachlorophenol	0.400	0.529	-32.3#	170	-0.03
23	Phenanthrene	0.400	0.393	1.8	107	0.00
24	Anthracene	0.400	0.392	2.0	113	0.00
25 SURR	Fluoranthene-d10	0.400	0.412	-3.0	112	0.00
26	Fluoranthene	0.400	0.411	-2.7	111	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
28	Pyrene	0.400	0.410	-2.5	115	0.00
29 S	Terphenyl-d14	0.400	0.415	-3.7	115	0.00
30	Benzo(a)anthracene	0.400	0.399	0.3	120	0.00
31	Chrysene	0.400	0.410	-2.5	110	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.511	-27.7#	164	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.412	-3.0	110	0.00
34 I	Perylene-d12	0.400	0.400	0.0	115	0.00
35	Benzo(b)fluoranthene	0.400	0.365	8.8	106	0.00
36	Benzo(k)fluoranthene	0.400	0.384	4.0	114	0.00
37 C	Benzo(a)pyrene	0.400	0.378	5.5	109	0.00
38	Dibenzo(a,h)anthracene	0.400	0.379	5.3	113	0.00
39	Benzo(g,h,i)perylene	0.400	0.384	4.0	112	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE062019\
Data File : BE100111.D
Acq On : 20 Jun 2019 17:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 21 02:06:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				