

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
 Data File : BE100113.D
 Acq On : 20 Jun 2019 18:50
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 21 02:07:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.642	0.652	-1.6	102	0.00
3	n-Nitrosodimethylamine	0.238	0.098	58.8#	124	0.00
4 S	2-Fluorophenol	0.830	0.825	0.6	93	0.00
5 S	Phenol-d6	1.140	1.098	3.7	85	0.00
6	bis(2-Chloroethyl)ether	1.119	1.006	10.1	89	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.388	0.356	8.2	80	0.01
9	Naphthalene	4.347	4.457	-2.5	90	0.00
10	Hexachlorobutadiene	0.427	0.431	-0.9	91	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.808	-6.2	97	0.00
12	2-Methylnaphthalene	0.715	0.765	-7.0	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.221	0.249	-12.7	139	0.00
15 S	2-Fluorobiphenyl	1.573	1.541	2.0	100	0.00
16	Acenaphthylene	1.899	1.884	0.8	108	-0.01
17	Acenaphthene	1.062	1.067	-0.5	108	0.00
18	Fluorene	1.581	1.695	-7.2	116	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.238	0.240	-0.8	125	0.00
21	Hexachlorobenzene	0.262	0.250	4.6	115	-0.01
22	Pentachlorophenol	0.053	0.062	-17.0	192#	-0.03
23	Phenanthrene	0.958	0.921	3.9	117	0.00
24	Anthracene	0.808	0.795	1.6	127	0.00
25 SURR	Fluoranthene-d10	5.718	5.804	-1.5	123	0.00
26	Fluoranthene	1.538	1.550	-0.8	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
28	Pvrene	1.028	1.073	-4.4	130	0.00
29 S	Terphenyl-d14	0.800	0.824	-3.0	126	0.00
30	Benzo(a)anthracene	1.190	1.254	-5.4	141	0.00
31	Chrysene	1.383	1.418	-2.5	122	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.455	-29.6#	184#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.628	-4.6	124	0.00
34 I	Perylene-d12	1.000	1.000	0.0	127	-0.01
35	Benzo(b)fluoranthene	1.174	1.117	4.9	122	0.00
36	Benzo(k)fluoranthene	1.335	1.240	7.1	122	0.00
37 C	Benzo(a)pyrene	1.133	1.139	-0.5	127	0.00
38	Dibenzo(a,h)anthracene	1.161	1.139	1.9	128	0.00
39	Benzo(a,h,i)perylene	1.207	1.168	3.2	124	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
Data File : BE100113.D
Acq On : 20 Jun 2019 18:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 21 02:07:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			