

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
 Data File : BE100129.D
 Acq On : 21 Jun 2019 4:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 21 06:52:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	-0.01
2	1,4-Dioxane	0.642	0.584	9.0	115	-0.01
3	n-Nitrosodimethylamine	0.238	0.197	17.2	314#	-0.07
4 S	2-Fluorophenol	0.830	1.567	-88.8#	223#	-0.01
5 S	Phenol-d6	1.140	2.265	-98.7#	219#	-0.01
6	bis(2-Chloroethyl)ether	1.119	1.282	-14.6	143	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.388	0.385	0.8	122	-0.01
9	Naphthalene	4.347	4.323	0.6	124	-0.01
10	Hexachlorobutadiene	0.427	0.383	10.3	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.815	-7.1	139	-0.01
12	2-Methylnaphthalene	0.715	0.825	-15.4	150	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	172#	-0.01
14 S	2,4,6-Tribromophenol	0.221	0.170	23.1	149	0.00
15 S	2-Fluorobiphenyl	1.573	1.460	7.2	148	-0.01
16	Acenaphthylene	1.899	1.930	-1.6	174#	-0.01
17	Acenaphthene	1.062	1.100	-3.6	175#	0.00
18	Fluorene	1.581	1.727	-9.2	185#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	181#	0.00
20	4-Bromophenyl-phenylether	0.238	0.255	-7.1	201#	-0.01
21	Hexachlorobenzene	0.262	0.260	0.8	181#	-0.01
22	Pentachlorophenol	0.053	0.011	79.2#	53	0.01
23	Phenanthrene	0.958	0.954	0.4	183#	0.00
24	Anthracene	0.808	0.867	-7.3	209#	0.00
25 SURR	Fluoranthene-d10	5.718	5.289	7.5	170#	0.00
26	Fluoranthene	1.538	1.411	8.3	168#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
28	Pvrene	1.028	1.209	-17.6	176#	0.00
29 S	Terphenyl-d14	0.800	0.934	-16.8	172#	-0.01
30	Benzo(a)anthracene	1.190	1.289	-8.3	175#	0.00
31	Chrysene	1.383	1.378	0.4	143	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.808	-61.0#	276#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.506	3.2	138	0.00
34 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
35	Benzo(b)fluoranthene	1.174	1.099	6.4	146	0.00
36	Benzo(k)fluoranthene	1.335	1.220	8.6	146	0.00
37 C	Benzo(a)pyrene	1.133	1.099	3.0	150	0.00
38	Dibenzo(a,h)anthracene	1.161	1.022	12.0	140	0.00
39	Benzo(a,h,i)perylene	1.207	1.040	13.8	134	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062019\
Data File : BE100129.D
Acq On : 21 Jun 2019 4:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 21 06:52:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			