

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
 Data File : BE100131.D
 Acq On : 21 Jun 2019 7:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 22 02:01:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.642	0.597	7.0	95	-0.01
3	n-Nitrosodimethylamine	0.238	0.345	-45.0#	443#	-0.04
4 S	2-Fluorophenol	0.830	1.330	-60.2#	153#	0.00
5 S	Phenol-d6	1.140	1.603	-40.6#	125	0.01
6	bis(2-Chloroethyl)ether	1.119	1.027	8.2	93	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.388	0.354	8.8	93	0.00
9	Naphthalene	4.347	4.360	-0.3	104	0.00
10	Hexachlorobutadiene	0.427	0.361	15.5	90	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.813	-6.8	115	0.00
12	2-Methylnaphthalene	0.715	0.803	-12.3	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	136	0.00
14 S	2,4,6-Tribromophenol	0.221	0.079	64.3#	55	0.01
15 S	2-Fluorobiphenyl	1.573	1.516	3.6	121	0.00
16	Acenaphthylene	1.899	1.969	-3.7	140	0.00
17	Acenaphthene	1.062	1.115	-5.0	140	0.00
18	Fluorene	1.581	1.683	-6.5	142	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	144	0.00
20	4-Bromophenyl-phenylether	0.238	0.240	-0.8	150#	0.00
21	Hexachlorobenzene	0.262	0.248	5.3	136	-0.01
22	Pentachlorophenol	0.053	0.009	83.0#	35#	0.07
23	Phenanthrene	0.958	0.945	1.4	144	0.00
24	Anthracene	0.808	0.784	3.0	150#	0.00
25 SURR	Fluoranthene-d10	5.718	5.394	5.7	137	0.00
26	Fluoranthene	1.538	1.434	6.8	136	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
28	Pvrene	1.028	1.168	-13.6	141	0.00
29 S	Terphenyl-d14	0.800	0.878	-9.7	135	0.00
30	Benzo(a)anthracene	1.190	1.266	-6.4	143	0.00
31	Chrysene	1.383	1.328	4.0	115	0.00
32	Bis(2-ethylhexyl)phthalate	1.123	1.459	-29.9#	185#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	0.810	47.9#	62	0.01
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.174	1.038	11.6	97	0.00
36	Benzo(k)fluoranthene	1.335	1.464	-9.7	123	0.00
37 C	Benzo(a)pyrene	1.133	1.076	5.0	103	0.00
38	Dibenzo(a,h)anthracene	1.161	0.570	50.9#	55	0.03
39	Benzo(a,h,i)perylene	1.207	0.721	40.3#	66	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
Data File : BE100131.D
Acq On : 21 Jun 2019 7:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 02:01:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			