

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
 Data File : BE100139.D
 Acq On : 21 Jun 2019 16:51
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 22 02:03:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.01
2	1,4-Dioxane	0.642	0.682	-6.2	110	-0.01
3	n-Nitrosodimethylamine	0.238	0.267	-12.2	349#	-0.04
4 S	2-Fluorophenol	0.830	1.178	-41.9#	138	-0.01
5 S	Phenol-d6	1.140	1.643	-44.1#	131	-0.01
6	bis(2-Chloroethyl)ether	1.119	1.138	-1.7	104	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	106	-0.01
8 S	Nitrobenzene-d5	0.388	0.392	-1.0	107	-0.01
9	Naphthalene	4.347	4.362	-0.3	107	-0.01
10	Hexachlorobutadiene	0.427	0.373	12.6	95	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.839	-10.2	122	-0.01
12	2-Methylnaphthalene	0.715	0.803	-12.3	125	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	136	-0.01
14 S	2,4,6-Tribromophenol	0.221	0.252	-14.0	174#	-0.01
15 S	2-Fluorobiphenyl	1.573	1.534	2.5	122	-0.01
16	Acenaphthylene	1.899	1.987	-4.6	141	-0.01
17	Acenaphthene	1.062	1.110	-4.5	139	0.00
18	Fluorene	1.581	1.742	-10.2	147	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	148	-0.01
20	4-Bromophenyl-phenylether	0.238	0.239	-0.4	155#	-0.01
21	Hexachlorobenzene	0.262	0.248	5.3	141	-0.01
22	Pentachlorophenol	0.053	0.056	-5.7	218#	-0.04
23	Phenanthrene	0.958	0.934	2.5	147	0.00
24	Anthracene	0.808	0.814	-0.7	161#	0.00
25 SURR	Fluoranthene-d10	5.718	5.519	3.5	145	0.00
26	Fluoranthene	1.538	1.461	5.0	143	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
28	Pvrene	1.028	1.269	-23.4	149	0.00
29 S	Terphenyl-d14	0.800	0.941	-17.6	140	-0.01
30	Benzo(a)anthracene	1.190	1.253	-5.3	137	0.00
31	Chrysene	1.383	1.388	-0.4	116	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	2.021	-80.0#	249#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.308	15.9	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	-0.01
35	Benzo(b)fluoranthene	1.174	1.095	6.7	113	0.00
36	Benzo(k)fluoranthene	1.335	1.190	10.9	111	-0.01
37 C	Benzo(a)pyrene	1.133	1.053	7.1	112	0.00
38	Dibenzo(a,h)anthracene	1.161	0.852	26.6#	91	0.00
39	Benzo(a,h,i)perylene	1.207	1.091	9.6	110	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
Data File : BE100139.D
Acq On : 21 Jun 2019 16:51
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 02:03:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			