

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
 Data File : BE100141.D
 Acq On : 21 Jun 2019 18:04
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 22 02:03:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.01
2	1,4-Dioxane	0.642	0.656	-2.2	106	-0.01
3	n-Nitrosodimethylamine	0.238	0.214	10.1	278#	-0.03
4 S	2-Fluorophenol	0.830	1.189	-43.3#	138	-0.01
5 S	Phenol-d6	1.140	1.560	-36.8#	124	-0.01
6	bis(2-Chloroethyl)ether	1.119	1.116	0.3	102	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.388	0.377	2.8	104	0.00
9	Naphthalene	4.347	4.340	0.2	108	-0.01
10	Hexachlorobutadiene	0.427	0.370	13.3	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.821	-7.9	121	0.00
12	2-Methylnaphthalene	0.715	0.799	-11.7	125	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	143	-0.01
14 S	2,4,6-Tribromophenol	0.221	0.183	17.2	133	0.00
15 S	2-Fluorobiphenyl	1.573	1.467	6.7	123	-0.01
16	Acenaphthylene	1.899	1.861	2.0	139	-0.01
17	Acenaphthene	1.062	1.049	1.2	138	0.00
18	Fluorene	1.581	1.635	-3.4	145	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.238	0.243	-2.1	151#	0.00
21	Hexachlorobenzene	0.262	0.260	0.8	142	-0.01
22	Pentachlorophenol	0.053	0.030	43.4#	111	-0.04
23	Phenanthrene	0.958	0.949	0.9	143	0.00
24	Anthracene	0.808	0.793	1.9	150#	0.00
25 SURR	Fluoranthene-d10	5.718	5.442	4.8	137	0.00
26	Fluoranthene	1.538	1.463	4.9	137	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.028	1.291	-25.6#	141	0.00
29 S	Terphenyl-d14	0.800	0.912	-14.0	127	0.00
30	Benzo(a)anthracene	1.190	1.123	5.6	114	0.00
31	Chrysene	1.383	1.389	-0.4	108	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	1.717	-52.9#	197#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.118	28.1#	77	0.00
34 I	Perylene-d12	1.000	1.000	0.0	106	0.00
35	Benzo(b)fluoranthene	1.174	1.013	13.7	92	0.00
36	Benzo(k)fluoranthene	1.335	1.193	10.6	98	0.00
37 C	Benzo(a)pyrene	1.133	1.041	8.1	98	0.00
38	Dibenzo(a,h)anthracene	1.161	0.654	43.7#	62	0.00
39	Benzo(a,h,i)perylene	1.207	1.058	12.3	94	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
Data File : BE100141.D
Acq On : 21 Jun 2019 18:04
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 02:03:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		