

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
 Data File : BE100154.D
 Acq On : 22 Jun 2019 4:00
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 22 04:40:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 19 15:07:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	169#	-0.01
2	1,4-Dioxane	0.642	0.409	36.3#	122	-0.01
3	n-Nitrosodimethylamine	0.238	0.265	-11.3	638#	-0.06
4 S	2-Fluorophenol	0.830	1.111	-33.9#	239#	-0.01
5 S	Phenol-d6	1.140	1.672	-46.7#	245#	-0.02
6	bis(2-Chloroethyl)ether	1.119	1.216	-8.7	205#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	190#	-0.01
8 S	Nitrobenzene-d5	0.388	0.422	-8.8	205#	-0.02
9	Naphthalene	4.347	4.409	-1.4	194#	-0.01
10	Hexachlorobutadiene	0.427	0.392	8.2	179#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.761	0.775	-1.8	202#	-0.01
12	2-Methylnaphthalene	0.715	0.771	-7.8	214#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	202#	-0.01
14 S	2,4,6-Tribromophenol	0.221	0.290	-31.2#	297#	-0.01
15 S	2-Fluorobiphenyl	1.573	1.639	-4.2	194#	-0.03
16	Acenaphthylene	1.899	2.061	-8.5	217#	-0.01
17	Acenaphthene	1.062	1.112	-4.7	207#	0.00
18	Fluorene	1.581	1.587	-0.4	199#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	175#	-0.01
20	4-Bromophenyl-phenylether	0.238	0.262	-10.1	200#	-0.01
21	Hexachlorobenzene	0.262	0.272	-3.8	182#	-0.01
22	Pentachlorophenol	0.053	0.103	-94.3#	466#	-0.05
23	Phenanthrene	0.958	0.985	-2.8	182#	-0.01
24	Anthracene	0.808	0.904	-11.9	210#	-0.01
25 SURR	Fluoranthene-d10	5.718	5.261	8.0	163#	0.00
26	Fluoranthene	1.538	1.406	8.6	162#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
28	Pvrene	1.028	1.264	-23.0	169#	-0.01
29 S	Terphenyl-d14	0.800	0.990	-23.7	168#	-0.02
30	Benzo(a)anthracene	1.190	1.354	-13.8	169#	0.00
31	Chrysene	1.383	1.359	1.7	130	-0.01
32	Bis(2-ethylhexyl)phthalate	1.123	2.283	-103.3#	320#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.630	-4.8	137	0.00
34 I	Perylene-d12	1.000	1.000	0.0	147	-0.01
35	Benzo(b)fluoranthene	1.174	1.095	6.7	138	0.00
36	Benzo(k)fluoranthene	1.335	1.153	13.6	131	0.00
37 C	Benzo(a)pyrene	1.133	1.058	6.6	137	-0.01
38	Dibenzo(a,h)anthracene	1.161	1.039	10.5	136	0.00
39	Benzo(a,h,i)perylene	1.207	1.067	11.6	131	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
Data File : BE100154.D
Acq On : 22 Jun 2019 4:00
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 22 04:40:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			