

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
 Data File : BE100153.D
 Acq On : 22 Jun 2019 3:25
 Operator : JU/SJ
 Sample : K3428-21
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 EXT-013-SW162-061319

Manual Integrations
 APPROVED

mohammad
 6/25/2019 3:21:47 PM

Quant Time: Jun 22 04:42:21 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Jun 19 15:07:24 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	869	0.40	ng	-0.01
7) Naphthalene-d8	10.52	136	3328	0.40	ng	-0.01
13) Acenaphthene-d10	14.40	164	2720	0.40	ng	-0.01
19) Phenanthrene-d10	17.15	188	7637	0.40	ng	-0.01
27) Chrysene-d12	21.34	240	10183	0.40	ng	0.00
34) Perylene-d12	23.84	264	11857	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.29	112	629	0.35	ng	-0.01
5) Phenol-d6	6.91	99	958	0.39	ng	-0.02
8) Nitrobenzene-d5	8.89	82	1039	0.32	ng	-0.03
11) 2-Methylnaphthalene-d10	12.14	152	147	0.02	ng	-0.01
14) 2,4,6-Tribromophenol	15.90	330	542	0.36	ng	-0.01
15) 2-Fluorobiphenyl	13.03	172	2721	0.25	ng	-0.01
25) Fluoranthene-d10	19.19	212	1595	0.01	ng	0.00
29) Terphenyl-d14	19.79	244	4370	0.21	ng	-0.02

Target Compounds

						Qvalue
9) Naphthalene	10.58	128	6421	0.178	ng	100
12) 2-Methylnaphthalene	12.21	142	1118	0.188	ng	97
16) Acenaphthylene	14.13	152	6739	0.522	ng	99
17) Acenaphthene	14.47	154	1219	0.169	ng	93
18) Fluorene	15.45	166	4063	0.378	ng	# 92
23) Phenanthrene	17.19	178	83918	4.586	ng	98
24) Anthracene	17.28	178	14434	0.935	ng	98
26) Fluoranthene	19.22	202	269849	9.187	ng	99
28) Pyrene	19.59	202	269913	10.313	ng	# 95
30) Benzo(a)anthracene	21.33	228	210467	6.948	ng	98
31) Chrysene	21.38	228	245822m	6.984	ng	
32) Bis(2-ethylhexyl)phthalate	21.25	149	149181	5.219	ng	99
33) Indeno(1,2,3-cd)pyrene	26.48	276	163167	4.120	ng	# 98
35) Benzo(b)fluoranthene	23.07	252	319903	9.194	ng	# 97
36) Benzo(k)fluoranthene	23.11	252	118548m	2.995	ng	
37) Benzo(a)pyrene	23.72	252	222756	6.632	ng	# 95
38) Dibenzo(a,h)anthracene	26.49	278	40247	1.169	ng	95
39) Benzo(a,h,i)perylene	27.30	276	166806	4.663	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062119\
Data File : BE100153.D
Acq On : 22 Jun 2019 3:25
Operator : JU/SJ
Sample : K3428-21
Misc :
ALS Vial : 14 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM

Quant Time: Jun 22 04:42:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 19 15:07:24 2019
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.72 min Scan# 422
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
152	100		
150	133.7	99.8	149.8
115	62.3	51.3	76.9

Manual Integrations
APPROVED




mohammad
6/25/2019 3:21:47 PM




#4

2-Fluorophenol
Concen: 0.349 ng
RT: 5.29 min Scan# 211
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
112	100		
64	55.6	22.0	33.0#
63	156.0	121.8	182.6





#5

Phenol-d6

Concen: 0.387 ng

RT: 6.91 min Scan# 351

Delta R.T. -0.02 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Instrument :

BNA_E





ClientSampleId :

EXT-013-SW162-061319

Tot Ion:	99	Resp:	958
Ion	Ratio	Lower	Upper
99	100		
42	14.9	0.0	0.0#
71	41.1	37.6	56.4

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM



#7

Naphthalene-d8

Concen: 0.400 ng




RT: 10.52 min Scan# 643


Delta R.T. -0.01 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Tgt Ion:	136	Resp:	3328
Ion	Ratio	Lower	Upper
136	100		
137	13.5	11.6	17.4
54	29.8	33.9	50.9#
68	7.5	7.8	11.8#





#8

Nitrobenzene-d5

Concen: 0.322 ng

RT: 8.89 min Scan# 517

Delta R.T. -0.03 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Instrument :

BNA_E


ClientSampleId :

EXT-013-SW162-061319




Tot Ion:	82	Resp:	1039
Ion Ratio	Lower	Upper	
82	100		
128	160.3	128.0	192.0
54	170.0	143.4	215.0

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM



Abundance Ion 82.00 (81.70 to 82.70): BE1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BE1



#9

Naphthalene

Concen: 0.178 ng

RT: 10.58 min Scan# 647




Delta R.T. -0.01 min


Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Tgt Ion:	128	Resp:	6421
Ion Ratio	Lower	Upper	
128	100		
129	4.1	3.1	4.7
127	4.3	3.4	5.2

Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#11
2-Methylnaphthalene-d10
Concen: 0.023 ng
RT: 12.14 min Scan# 758
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion:152 Resp: 147
Ion Ratio Lower Upper
152 100
151 23.8 16.3 24.5




Manual Integrations
APPROVED


mohammad
6/25/2019 3:21:47 PM



#12
2-Methylnaphthalene
Concen: 0.188 ng
RT: 12.21 min Scan# 763
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion:142 Resp: 1118
Ion Ratio Lower Upper
142 100
141 90.6 74.1 111.1
115 47.0 40.9 61.3






#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.40 min Scan# 915
 Delta R.T. -0.01 min
 Lab File: BE100153.D
 Acq: 22 Jun 2019 3:25

Instrument :
 BNA_E
 ClientSampleId :
 EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
164	100		
162	98.1	78.2	117.2
160	51.4	45.6	68.4

Manual Integrations
 APPROVED

mohammad
 6/25/2019 3:21:47 PM






Abundance

Ion 164.00 (163.70 to 164.70): E


Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14
 2,4,6-Tribromophenol
 Concen: 0.361 ng
 RT: 15.90 min Scan# 1021
 Delta R.T. -0.01 min
 Lab File: BE100153.D
 Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
330	100		
332	103.5	73.3	109.9
141	59.4	22.2	33.2#






Abundance

Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E






#15
2-Fluorobiphenyl
Concen: 0.254 ng
RT: 13.03 min Scan# 820
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.1	29.8	44.8
170	25.6	23.2	34.8

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM






Abundance

Ion 172.00 (171.70 to 172.70): E


Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#16
Acenaphthylene
Concen: 0.522 ng
RT: 14.13 min Scan# 896
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.9	23.9
153	14.0	10.1	15.1






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E






#17
Acenaphthene
Concen: 0.169 ng
RT: 14.47 min Scan# 920
Delta R.T. -0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
154	100		
153	126.4	95.5	143.3
152	62.9	54.9	82.3

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM






Abundance

Ion 154.00 (153.70 to 154.70): E


Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18
Fluorene
Concen: 0.378 ng
RT: 15.45 min Scan# 987
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
166	100		
165	92.1	79.8	119.8
167	16.4	10.2	15.4#






Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.15 min Scan# 1114

Delta R.T. -0.01 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Instrument :

BNA_E





ClientSampleId :

EXT-013-SW162-061319

Tot Ion:	188	Resp:	7637
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	8.8	6.3	9.5

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM



#23

Phenanthrene

Concen: 4.586 ng




RT: 17.19 min Scan# 1117


Delta R.T. -0.01 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Tgt Ion:	178	Resp:	83918
Ion Ratio	Lower	Upper	
178	100		
176	19.6	15.9	23.9
179	15.3	13.1	19.7









#24
 Anthracene
 Concen: 0.935 ng
 RT: 17.28 min Scan# 1124
 Delta R.T. -0.01 min
 Lab File: BE100153.D
 Acq: 22 Jun 2019 3:25

Instrument :
 BNA_E
 ClientSampleId :
 EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.8	23.6
179	13.9	12.0	18.0




Manual Integrations
 APPROVED


mohammad
 6/25/2019 3:21:47 PM



#25
 Fluoranthene-d10
 Concen: 0.015 ng
 RT: 19.19 min Scan# 1379
 Delta R.T. -0.01 min
 Lab File: BE100153.D
 Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
212	100		
106	3.4	1.4	2.0#
104	2.1	0.8	1.2#









#26
Fluoranthene
Concen: 9.187 ng
RT: 19.22 min Scan# 1384
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
202	100		
101	5.7	5.0	7.4
203	17.6	13.7	20.5




Manual Integrations
APPROVED


mohammad
6/25/2019 3:21:47 PM



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.34 min Scan# 1628
Delta R.T. 0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.1	6.6	10.0#
236	30.1	23.8	35.6









#28
Pvrene
Concen: 10.313 ng
RT: 19.59 min Scan# 1448
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Resp Lower	Upper
202	100		
200	19.8	15.6	23.4
203	21.8	13.9	20.9




Manual Integrations
APPROVED


mohammad
6/25/2019 3:21:47 PM



#29
Terphenyl-d14
Concen: 0.215 ng
RT: 19.79 min Scan# 1483
Delta R.T. -0.02 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Resp Lower	Upper
244	100		
212	44.2	29.8	44.8
122	8.4	6.1	9.1






#30
Benzo(a)anthracene
Concen: 6.948 ng
RT: 21.33 min Scan# 1627
Delta R.T. 0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
228	100		
226	28.4	23.0	34.4
229	22.0	16.4	24.6

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM




Abundance



Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



Time-->



#31
Chrysene
Concen: 6.984 ng m
RT: 21.38 min Scan# 1631
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.5	23.8	35.8
229	21.5	16.3	24.5

Abundance



Ion 228.00 (227.70 to 228.70): E


Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



Time-->









#32
Bis(2-ethylhexyl)phthalate
Concen: 5.219 ng
RT: 21.25 min Scan# 1620
Delta R.T. 0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
149	100		
167	7.4	5.8	8.6
279	1.6	1.1	1.7




Manual Integrations
APPROVED


mohammad
6/25/2019 3:21:47 PM



#33
Indeno(1,2,3-cd)pyrene
Concen: 4.120 ng
RT: 26.48 min Scan# 2927
Delta R.T. -0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
276	100		
138	9.1	4.3	6.5#
277	24.0	19.4	29.0






#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.84 min Scan# 2191
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
264	100		
260	27.7	20.8	31.2
265	41.6	23.2	34.8#

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#35
Benzo(b)fluoranthene
Concen: 9.194 ng
RT: 23.07 min Scan# 1976
Delta R.T. -0.00 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.4	27.6
125	5.1	6.4	9.6#


Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#36
Benzo(k)fluoranthene
Concen: 2.995 ng m
RT: 23.11 min Scan# 1988
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Instrument :
BNA_E
ClientSampleId :
EXT-013-SW162-061319

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.5	18.9	28.3
125	5.5	5.3	7.9




Manual Integrations
APPROVED


mohammad
6/25/2019 3:21:47 PM



#37
Benzo(a)pyrene
Concen: 6.632 ng
RT: 23.72 min Scan# 2159
Delta R.T. -0.01 min
Lab File: BE100153.D
Acq: 22 Jun 2019 3:25

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.2	19.4	29.2
125	5.2	8.0	12.0





#38

Dibenzo(a,h)anthracene

Concen: 1.169 ng

RT: 26.49 min Scan# 2930

Delta R.T. -0.02 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Instrument :

BNA_E


ClientSampleId :

EXT-013-SW162-061319

Tot Ion:	278	Resp:	40247
Ion Ratio	Lower	Upper	
278	100		
139	9.1	9.0	13.6
279	29.5	21.4	32.0

Manual Integrations
APPROVED

mohammad
6/25/2019 3:21:47 PM






Abundance

Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 4.663 ng

RT: 27.30 min Scan# 3157

Delta R.T. 0.00 min

Lab File: BE100153.D

Acq: 22 Jun 2019 3:25

Tgt Ion:	276	Resp:	166806
Ion Ratio	Lower	Upper	
276	100		
277	24.6	20.3	30.5
138	9.7	11.3	16.9#

Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

